

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070324\
 Data File : VU059766.D
 Acq On : 03 Jul 2024 02:44
 Operator : MD/SY
 Sample : P3121-20DL 40X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COBM8DL

Quant Time: Jul 03 04:11:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 03 04:07:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	169308	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	171638	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	79350	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	54958	4.668	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	93.400%	
7) Chloroethane-d5	1.901	69	60781	5.366	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	107.400%	
11) 1,1-Dichloroethene-d2	2.554	65	22117	4.528	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	90.600%	
20) 2-Butanone-d5	4.621	46	111026	45.783	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	91.560%	
24) Chloroform-d	5.052	84	99371	4.112	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	82.200%	
26) 1,2-Dichloroethane-d4	5.692	65	49303	4.236	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.800%	
32) Benzene-d6	5.717	84	207302	4.388	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	87.800%	
36) 1,2-Dichloropropane-d6	6.682	67	64612	4.464	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	89.200%	
41) Toluene-d8	7.891	98	176704	4.024	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	80.400%	
43) trans-1,3-Dichloroprop...	8.174	79	20741	4.576	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	91.600%	
46) 2-Hexanone-d5	8.630	63	94777	50.326	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	100.660%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	52324	4.375	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	87.600%	
66) 1,2-Dichlorobenzene-d4	12.187	152	64899	4.595	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	92.000%	
Target Compounds						
						Qvalue
16) Methylene chloride	3.036	84	6542	0.460	ug/L	87
22) cis-1,2-Dichloroethene	4.656	96	4760	0.346	ug/L	87
47) Tetrachloroethene	8.547	164	77971	7.118	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070324\
Data File : VU059766.D
Acq On : 03 Jul 2024 02:44
Operator : MD/SY
Sample : P3121-20DL 40X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BM8DL

Quant Time: Jul 03 04:11:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 03 04:07:12 2024
Response via : Initial Calibration

