

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070324\
 Data File : VU059793.D
 Acq On : 03 Jul 2024 13:20
 Operator : MD/SY
 Sample : P3109-03 50X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 74 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H0HC3

Quant Time: Jul 04 00:58:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jul 03 06:10:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	169601	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	172641	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	77172	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	51245	4.345	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.000%
7) Chloroethane-d5	1.907	69	42984	3.788	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.800%
11) 1,1-Dichloroethene-d2	2.560	65	19169	3.918	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.400%
20) 2-Butanone-d5	4.621	46	97188	40.008	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	80.020%
24) Chloroform-d	5.055	84	91461	3.778	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	75.600%
26) 1,2-Dichloroethane-d4	5.695	65	44992	3.859	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.200%
32) Benzene-d6	5.721	84	187344	3.943	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.800%
36) 1,2-Dichloropropane-d6	6.682	67	58754	4.036	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.800%
41) Toluene-d8	7.894	98	161996	3.668	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.400%
43) trans-1,3-Dichloroprop...	8.177	79	20376	4.469	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.400%
46) 2-Hexanone-d5	8.627	63	78558	41.471	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.940%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	46324	3.851	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	77.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	58847	4.284	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.600%
Target Compounds						
					Qvalue	
16) Methylene chloride	3.039	84	9442	0.662	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070324\
Data File : VU059793.D
Acq On : 03 Jul 2024 13:20
Operator : MD/SY
Sample : P3109-03 50X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 74 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H0HC3

Quant Time: Jul 04 00:58:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jul 03 06:10:28 2024
Response via : Initial Calibration

