

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070324\
 Data File : VU059798.D
 Acq On : 03 Jul 2024 15:42
 Operator : MD/SY
 Sample : VU0703WBL04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 80 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK012

Quant Time: Jul 04 02:09:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jul 04 02:08:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	158101	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	168675	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	69104	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	58237	5.297	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	106.000%	
7) Chloroethane-d5	1.907	69	48290	4.565	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	91.400%	
11) 1,1-Dichloroethene-d2	2.557	65	21396	4.691	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	93.800%	
20) 2-Butanone-d5	4.618	46	105689	46.672	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	93.340%	
24) Chloroform-d	5.055	84	102726	4.552	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.000%	
26) 1,2-Dichloroethane-d4	5.695	65	50207	4.619	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	92.400%	
32) Benzene-d6	5.721	84	208838	4.499	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.000%	
36) 1,2-Dichloropropane-d6	6.682	67	65265	4.588	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	91.800%	
41) Toluene-d8	7.894	98	176538	4.091	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	81.800%	
43) trans-1,3-Dichloroprop...	8.177	79	21895	4.915	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	98.400%	
46) 2-Hexanone-d5	8.627	63	81482	44.026	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	88.060%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	52568	4.473	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	89.400%	
66) 1,2-Dichlorobenzene-d4	12.187	152	64944	5.280	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.600%	
Target Compounds						
					Qvalue	
16) Methylene chloride	3.039	84	2765	0.208	ug/L	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070324\
Data File : VU059798.D
Acq On : 03 Jul 2024 15:42
Operator : MD/SY
Sample : VU0703WBL04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 80 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK012

Quant Time: Jul 04 02:09:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jul 04 02:08:13 2024
Response via : Initial Calibration

