

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070419\
 Data File : VU033178.D
 Acq On : 03 Jul 2019 18:05
 Operator : JC/SP
 Sample : K3598-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Jul 04 02:23:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 04 01:20:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	77417	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.08	117	80474	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	35718	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	22595	5.04	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.80%
7) Chloroethane-d5	1.68	69	20951	5.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.20%
11) 1,1-Dichloroethene-d2	2.27	63	37340	3.54	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.80%
20) 2-Butanone-d5	4.19	46	90183	55.28	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.56%
24) Chloroform-d	4.64	84	53663	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.00%
26) 1,2-Dichloroethane-d4	5.30	65	32803	5.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.00%
32) Benzene-d6	5.33	84	95616	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
36) 1,2-Dichloropropane-d6	6.32	67	33524	5.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.40%
41) Toluene-d8	7.56	98	81734	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.00%
43) trans-1,3-Dichloropropene-	7.85	79	12241	4.43	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.60%
46) 2-Hexanone-d5	8.31	63	53311	46.80	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.60%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	27804	4.93	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.60%
64) 1,2-Dichlorobenzene-d4	11.85	152	34495	5.36	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070419\
Data File : VU033178.D
Acq On : 03 Jul 2019 18:05
Operator : JC/SP
Sample : K3598-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jul 04 02:23:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 04 01:20:02 2019
Response via : Initial Calibration

