

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070518\
 Data File : VU025208.D
 Acq On : 05 Jul 2018 12:10
 Operator : MD/SY
 Sample : J3808-18
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1TB04-180628

Quant Time: Jul 06 12:49:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	208668	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	311991	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	287876	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	147492	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	157202	46.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.30%	
35) Dibromofluoromethane	4.89	113	123573	47.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.44%	
50) Toluene-d8	7.57	98	446846	47.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.62%	
62) 4-Bromofluorobenzene	10.31	95	164495	43.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.74%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070518\
Data File : VU025208.D
Acq On : 05 Jul 2018 12:10
Operator : MD/SY
Sample : J3808-18
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1TB04-180628

Quant Time: Jul 06 12:49:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

