

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070518\
 Data File : VU025211.D
 Acq On : 05 Jul 2018 13:23
 Operator : MD/SY
 Sample : J3808-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021MW012-180627

Quant Time: Jul 06 05:43:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	209538	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	311303	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	288477	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	144646	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	153496	44.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.74%	
35) Dibromofluoromethane	4.89	113	121347	46.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.94%	
50) Toluene-d8	7.57	98	445066	47.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.46%	
62) 4-Bromofluorobenzene	10.31	95	162207	43.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.70%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.33	50	1338	0.499	ug/l	96
16) Acetone	2.32	43	5146	2.800	ug/l	99
27) cis-1,2-Dichloroethene	4.24	96	1920	0.636	ug/l	78
44) Trichloroethene	6.19	130	21495	7.246	ug/l	93
93) 1,2,4-Trichlorobenzene	13.52	180	1081	0.336	ug/l #	82
95) Naphthalene	13.75	128	8104	1.431	ug/l	96
96) 1,2,3-Trichlorobenzene	13.99	180	1630	0.509	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070518\
Data File : VU025211.D
Acq On : 05 Jul 2018 13:23
Operator : MD/SY
Sample : J3808-01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021MW012-180627

Quant Time: Jul 06 05:43:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

