

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070522\
 Data File : VU049608.D
 Acq On : 05 Jul 2022 11:11
 Operator : SY/MD
 Sample : VU0705WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0705WBL01

Quant Time: Jul 06 07:46:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	263530	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	454703	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	409816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	186558	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	174781	43.065	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.140%
35) Dibromofluoromethane	5.288	113	142717	44.862	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.720%
50) Toluene-d8	7.896	98	543186	47.872	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.740%
62) 4-Bromofluorobenzene	10.632	95	197472	45.432	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.860%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070522\
Data File : VU049608.D
Acq On : 05 Jul 2022 11:11
Operator : SY/MD
Sample : VU0705WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0705WBL01

Quant Time: Jul 06 07:46:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

