

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU070522\  
 Data File : VU049609.D  
 Acq On : 05 Jul 2022 11:35  
 Operator : SY/MD  
 Sample : VU0705MBL01  
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA\_U/MEOH  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VU0705MBL01

Quant Time: Jul 06 07:46:32 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U062722W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jun 28 10:35:43 2022  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 5.372  | 168  | 263136   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.247  | 114  | 452280   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 9.417  | 117  | 410192   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 11.812 | 152  | 188252   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |         |
|-----------------------------|--------|-------|----------|----------|------|---------|
| System Monitoring Compounds |        |       |          |          |      |         |
| 33) 1,2-Dichloroethane-d4   | 5.700  | 65    | 175825   | 43.387   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =    | 86.780% |
| 35) Dibromofluoromethane    | 5.288  | 113   | 142090   | 44.905   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =    | 89.800% |
| 50) Toluene-d8              | 7.896  | 98    | 546059   | 48.383   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =    | 96.760% |
| 62) 4-Bromofluorobenzene    | 10.632 | 95    | 197696   | 45.728   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =    | 91.460% |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU070522\  
Data File : VU049609.D  
Acq On : 05 Jul 2022 11:35  
Operator : SY/MD  
Sample : VU0705MBL01  
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA\_U/MEOH  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VU0705MBL01

Quant Time: Jul 06 07:46:32 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\82U062722W.M  
Quant Title : SW846 8260  
QLast Update : Tue Jun 28 10:35:43 2022  
Response via : Initial Calibration

