

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070522\
 Data File : VU049612.D
 Acq On : 05 Jul 2022 12:49
 Operator : SY/MD
 Sample : N3569-17DL 20X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 152140-SPC-09-80DL

Quant Time: Jul 06 07:47:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	273788	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	467458	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	434281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	197212	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	181962	43.155	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.300%
35) Dibromofluoromethane	5.289	113	148452	45.392	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.780%
50) Toluene-d8	7.896	98	567161	48.621	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.240%
62) 4-Bromofluorobenzene	10.632	95	208862	46.742	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.480%
Target Compounds						
64) Tetrachloroethene	8.552	164	139082	49.932	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070522\
Data File : VU049612.D
Acq On : 05 Jul 2022 12:49
Operator : SY/MD
Sample : N3569-17DL 20X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
152140-SPC-09-80DL

Quant Time: Jul 06 07:47:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

