

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070522\
 Data File : VU049616.D
 Acq On : 05 Jul 2022 14:24
 Operator : SY/MD
 Sample : N3564-30
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-16-2-4-GR

Quant Time: Jul 06 07:47:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	269678	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	474146	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	425857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	191730	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	182262	43.885	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.760%
35) Dibromofluoromethane	5.288	113	147237	44.385	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.780%
50) Toluene-d8	7.896	98	563168	47.598	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.200%
62) 4-Bromofluorobenzene	10.632	95	201033	44.355	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.720%
Target Compounds						
					Qvalue	
16) Acetone	2.626	43	41130	19.638	ug/l	98
20) Methylene Chloride	3.044	84	144389	41.284	ug/l	91
37) Ethyl Acetate	4.806	43	42314	6.438	ug/l #	94
43) Isopropyl Acetate	5.909	43	172921	17.955	ug/l	97
65) Chlorobenzene	9.446	112	232449	26.700	ug/l	99
95) Naphthalene	14.089	128	9010	2.315	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070522\
Data File : VU049616.D
Acq On : 05 Jul 2022 14:24
Operator : SY/MD
Sample : N3564-30
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-16-2-4-GR

Quant Time: Jul 06 07:47:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

