

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070522\
 Data File : VU049618.D
 Acq On : 05 Jul 2022 15:11
 Operator : SY/MD
 Sample : N3564-35
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-16-2-3-GR

Quant Time: Jul 06 07:48:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	274913	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	476054	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	428111	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	201289	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	182944	43.210	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.420%
35) Dibromofluoromethane	5.288	113	149842	44.990	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.980%
50) Toluene-d8	7.896	98	571308	48.093	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.180%
62) 4-Bromofluorobenzene	10.632	95	208629	45.847	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.700%
Target Compounds						
						Qvalue
16) Acetone	2.623	43	50848	23.815	ug/l	99
20) Methylene Chloride	3.044	84	50197	13.582	ug/l	88
31) Cyclohexane	5.378	56	15841	2.492	ug/l	# 58
37) Ethyl Acetate	4.803	43	45973	6.967	ug/l	# 93
39) Methylcyclohexane	6.761	83	13418	2.318	ug/l	92
43) Isopropyl Acetate	5.909	43	180094	18.625	ug/l	98
69) o-Xylene	10.099	106	9819	1.664	ug/l	88
73) Isopropylbenzene	10.481	105	24954	1.740	ug/l	95
78) n-propylbenzene	10.906	91	27480	1.577	ug/l	96
80) 1,3,5-Trimethylbenzene	11.089	105	24242	1.937	ug/l	95
84) 1,2,4-Trimethylbenzene	11.468	105	101813	8.279	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070522\
Data File : VU049618.D
Acq On : 05 Jul 2022 15:11
Operator : SY/MD
Sample : N3564-35
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-16-2-3-GR

Quant Time: Jul 06 07:48:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

