

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070522\
 Data File : VU049621.D
 Acq On : 05 Jul 2022 16:23
 Operator : SY/MD
 Sample : N3564-21 10X
 Misc : 5.77g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-16-2-5

Quant Time: Jul 06 07:48:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	268348	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	466431	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	427813	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	207708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	180588	43.697	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.400%
35) Dibromofluoromethane	5.289	113	146777	44.979	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.960%
50) Toluene-d8	7.896	98	563788	48.439	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.880%
62) 4-Bromofluorobenzene	10.632	95	216596	48.579	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070522\
Data File : VU049621.D
Acq On : 05 Jul 2022 16:23
Operator : SY/MD
Sample : N3564-21 10X
Misc : 5.77g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-16-2-5

Quant Time: Jul 06 07:48:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

