

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070522\
 Data File : VU049622.D
 Acq On : 05 Jul 2022 16:46
 Operator : SY/MD
 Sample : N3564-16 10X
 Misc : 4.87g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WC-16-2-6

Quant Time: Jul 06 07:48:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	273651	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	477436	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	437718	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	212909	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	184478	43.773	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.540%
35) Dibromofluoromethane	5.288	113	149505	44.758	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.520%
50) Toluene-d8	7.896	98	573316	48.122	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.240%
62) 4-Bromofluorobenzene	10.632	95	216265	47.387	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.780%
Target Compounds						
					Qvalue	
18) Methyl Acetate	2.948	43	10312	1.792	ug/l	92
52) Toluene	7.967	92	33591	3.865	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070522\
Data File : VU049622.D
Acq On : 05 Jul 2022 16:46
Operator : SY/MD
Sample : N3564-16 10X
Misc : 4.87g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WC-16-2-6

Quant Time: Jul 06 07:48:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration

