

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070522\
 Data File : VU049608.D
 Acq On : 05 Jul 2022 11:11
 Operator : SY/MD
 Sample : VU0705WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0705WBL01

Quant Time: Jul 06 07:46:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 28 10:35:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.372	168	263530	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	454703	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.417	117	409816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	186558	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.700	65	174781	43.065	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.140%
35) Dibromofluoromethane	5.288	113	142717	44.862	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.720%
50) Toluene-d8	7.896	98	543186	47.872	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.740%
62) 4-Bromofluorobenzene	10.632	95	197472	45.432	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.860%


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070522\
Data File : VU049608.D
Acq On : 05 Jul 2022 11:11
Operator : SY/MD
Sample : VU0705WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0705WBL01

Quant Time: Jul 06 07:46:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U062722W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 28 10:35:43 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.372 min Scan# 1254
Delta R.T. 0.003 min
Lab File: VU049608.D
Acq: 05 Jul 2022 11:11




Instrument :
MSVOA_U
ClientSampleId :
VU0705WBL01


Tgt Ion:168 Resp: 263530
Ion Ratio Lower Upper
168 100
99 59.8 39.3 58.9#



#33
1,2-Dichloroethane-d4
Concen: 43.065 ug/l
RT: 5.700 min Scan# 1356
Delta R.T. 0.000 min
Lab File: VU049608.D
Acq: 05 Jul 2022 11:11

Tgt Ion: 65 Resp: 174781
Ion Ratio Lower Upper
65 100
67 54.4 0.0 108.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.247 min Scan# 1

Delta R.T. 0.000 min

Lab File: VU049608.D

Acq: 05 Jul 2022 11:11

Instrument :

MSVOA_U

ClientSampleId :

VU0705WBL01





Tgt Ion:114 Resp: 454703

Ion Ratio Lower Upper

114 100

63 22.0 0.0 37.8

88 16.8 0.0 29.2



#35

Dibromofluoromethane

Concen: 44.862 ug/l

RT: 5.288 min Scan# 1228

Delta R.T. 0.000 min

Lab File: VU049608.D

Acq: 05 Jul 2022 11:11




Tgt Ion:113 Resp: 142717


Ion Ratio Lower Upper

113 100

111 103.6 82.1 123.1

192 19.0 18.0 27.0









#50
Toluene-d8
Concen: 47.872 ug/l
RT: 7.896 min Scan# 2039
Delta R.T. 0.000 min
Lab File: VU049608.D
Acq: 05 Jul 2022 11:11




Instrument :
MSVOA_U
ClientSampleId :
VU0705WBL01


Tgt Ion: 98 Resp: 543186
Ion Ratio Lower Upper
98 100
100 64.6 51.7 77.5



#62
4-Bromofluorobenzene
Concen: 45.432 ug/l
RT: 10.632 min Scan# 2890
Delta R.T. -0.003 min
Lab File: VU049608.D
Acq: 05 Jul 2022 11:11

Tgt Ion: 95 Resp: 197472
Ion Ratio Lower Upper
95 100
174 75.5 0.0 182.0
176 72.0 0.0 177.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.417 min Scan# 21
Delta R.T. -0.003 min
Lab File: VU049608.D
Acq: 05 Jul 2022 11:11

Instrument :
MSVOA_U
ClientSampleId :
VU0705WBL01

Tgt Ion:117 Resp: 409816
Ion Ratio Lower Upper
117 100
82 58.2 41.7 62.5
119 31.7 25.7 38.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 11.812 min Scan# 3257
Delta R.T. 0.000 min
Lab File: VU049608.D
Acq: 05 Jul 2022 11:11

Tgt Ion:152 Resp: 186558
Ion Ratio Lower Upper
152 100
115 61.8 37.0 110.9
150 155.5 0.0 342.2

