

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070523\
 Data File : VU054761.D
 Acq On : 05 Jul 2023 11:07
 Operator : MD/SY
 Sample : 03505-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Jul 06 01:02:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jul 06 01:01:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	218419	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	214603	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	96146	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	57096	4.107	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.200%
7) Chloroethane-d5	1.907	69	54712	4.204	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.000%
11) 1,1-Dichloroethene-d2	2.560	65	25858	4.165	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.400%
20) 2-Butanone-d5	4.624	46	224194	57.365	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.720%
24) Chloroform-d	5.058	84	133926	4.609	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.200%
26) 1,2-Dichloroethane-d4	5.698	65	71603	5.055	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.200%
32) Benzene-d6	5.721	84	242477	4.065	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.200%
36) 1,2-Dichloropropane-d6	6.685	67	87064	4.370	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.400%
41) Toluene-d8	7.894	98	202415	3.823	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.400%
43) trans-1,3-Dichloroprop...	8.177	79	31284	4.516	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.400%
46) 2-Hexanone-d5	8.631	63	152155	52.349	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.700%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	73996	4.981	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	78670	4.851	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.000%
Target Compounds						Qvalue
16) Methylene chloride	3.036	84	14989	0.673	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070523\
Data File : VU054761.D
Acq On : 05 Jul 2023 11:07
Operator : MD/SY
Sample : 03505-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Jul 06 01:02:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jul 06 01:01:36 2023
Response via : Initial Calibration

