

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070523\
 Data File : VU054762.D
 Acq On : 05 Jul 2023 11:31
 Operator : MD/SY
 Sample : 03495-09DL 10X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H0FB3DL

Quant Time: Jul 06 01:02:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jul 06 01:01:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	228812	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	223473	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	100575	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	54847	3.766	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	75.400%	
7) Chloroethane-d5	1.907	69	52887	3.879	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	77.600%	
11) 1,1-Dichloroethene-d2	2.560	65	22790	3.504	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	70.000%	
20) 2-Butanone-d5	4.624	46	210479	51.409	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	102.820%	
24) Chloroform-d	5.058	84	125277	4.116	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	82.400%	
26) 1,2-Dichloroethane-d4	5.698	65	67599	4.556	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	91.200%	
32) Benzene-d6	5.724	84	229295	3.691	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	73.800%	
36) 1,2-Dichloropropane-d6	6.685	67	81437	3.925	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	78.600%	
41) Toluene-d8	7.897	98	192910	3.499	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	70.000%	
43) trans-1,3-Dichloroprop...	8.177	79	30368	4.209	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	84.200%	
46) 2-Hexanone-d5	8.631	63	143885	47.539	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	95.080%	
56) 1,1,2,2-Tetrachloroeth...	10.753	84	70309	4.545	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	90.800%	
66) 1,2-Dichlorobenzene-d4	12.193	152	75986	4.479	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	89.600%	
Target Compounds						
					Qvalue	
16) Methylene chloride	3.042	84	16433	0.704	ug/L	96
34) Trichloroethene	6.541	95	11360	0.535	ug/L	98
47) Tetrachloroethene	8.550	164	127170	7.822	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070523\
Data File : VU054762.D
Acq On : 05 Jul 2023 11:31
Operator : MD/SY
Sample : 03495-09DL 10X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H0FB3DL

Quant Time: Jul 06 01:02:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR060723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jul 06 01:01:36 2023
Response via : Initial Calibration

