

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070620\
 Data File : VU039289.D
 Acq On : 06 Jul 2020 14:28
 Operator : SY/MD
 Sample : L3193-07MS
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H4296MS

Quant Time: Jul 07 01:46:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jul 07 01:17:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	120716	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	111976	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	53663	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	46945	5.06	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.20%
7) Chloroethane-d5	1.93	69	40191	4.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.60%
11) 1,1-Dichloroethene-d2	2.58	63	88669	5.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	102.80%
20) 2-Butanone-d5	4.66	46	163204	46.64	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.28%
24) Chloroform-d	5.08	84	87842	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
26) 1,2-Dichloroethane-d4	5.72	65	52000	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.60%
32) Benzene-d6	5.74	84	174638	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.60%
36) 1,2-Dichloropropane-d6	6.71	67	58488	5.11	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.20%
41) Toluene-d8	7.91	98	160038	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.00%
43) trans-1,3-Dichloropropene-	8.19	79	22262	4.20	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.00%
46) 2-Hexanone-d5	8.64	63	135971	48.72	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.44%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	51944	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	60568	5.00	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.00%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl chloride	1.61	62	38094	4.236	ug/L #	64
12) 1,1-Dichloroethene	2.59	96	35650	5.213	ug/L	95
16) Methylene chloride	3.06	84	1334	0.153	ug/L #	76
18) trans-1,2-Dichloroethene	3.37	96	6456	0.868	ug/L	87
22) cis-1,2-Dichloroethene	4.69	96	12286	1.450	ug/L	98
30) Cyclohexane	5.40	56	10983	0.853	ug/L	95
33) Benzene	5.79	78	173684	5.503	ug/L	100
34) Trichloroethene	6.56	95	43005	5.285	ug/L	93
42) Toluene	7.98	91	182721	5.319	ug/L	99
51) Chlorobenzene	9.45	112	110335	5.194	ug/L	98
53) m,p-Xylene	9.70	106	1552	0.108	ug/L	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070620\
Data File : VU039289.D
Acq On : 06 Jul 2020 14:28
Operator : SY/MD
Sample : L3193-07MS
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4296MS

Quant Time: Jul 07 01:46:14 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 07 01:17:04 2020
Response via : Initial Calibration

