

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070620\
 Data File : VU039293.D
 Acq On : 06 Jul 2020 16:01
 Operator : SY/MD
 Sample : L3193-03
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H4279

Quant Time: Jul 07 02:04:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jul 07 01:17:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.27	114	116231	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	109381	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	51372	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	47660	5.33	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.60%
7) Chloroethane-d5	1.93	69	39258	4.59	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.80%
11) 1,1-Dichloroethene-d2	2.58	63	65639	3.95	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.00%
20) 2-Butanone-d5	4.66	46	179349	53.24	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.48%
24) Chloroform-d	5.09	84	81920	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
26) 1,2-Dichloroethane-d4	5.72	65	50003	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
32) Benzene-d6	5.74	84	167314	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.60%
36) 1,2-Dichloropropane-d6	6.70	67	55597	4.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.40%
41) Toluene-d8	7.91	98	153561	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.20%
43) trans-1,3-Dichloropropene-	8.19	79	21032	4.07	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.40%
46) 2-Hexanone-d5	8.64	63	135213	49.60	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.20%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	50342	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	58109	5.01	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.20%

Target Compounds


					Ovalue
5) Vinyl chloride	1.61	62	11268	1.301 ug/L #	1
17) Methyl tert-butyl Ether	3.38	73	11472	0.530 ug/L #	95
18) trans-1,2-Dichloroethene	3.37	96	1008	0.141 ug/L	86
19) 1,1-Dichloroethane	3.90	63	4129	0.288 ug/L	98
22) cis-1,2-Dichloroethene	4.69	96	16414	2.012 ug/L	93
33) Benzene	5.79	78	4172	0.135 ug/L	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070620\
Data File : VU039293.D
Acq On : 06 Jul 2020 16:01
Operator : SY/MD
Sample : L3193-03
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4279

Quant Time: Jul 07 02:04:18 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jul 07 01:17:04 2020
Response via : Initial Calibration





#5

Vinyl chloride

Concen: 1.301 ug/L

RT: 1.61 min Scan# 84

Delta R.T. -0.00 min

Lab File: VU039293.D

Acq: 06 Jul 2020 16:01

Instrument :

MSVOA_U

ClientSampleId :





H4279

Tot Ion: 62 Resp: 11268

Ion Ratio Lower Upper

62 100

64 162.5 25.6 47.4#



#17

Methyl tert-butyl Ether

Concen: 0.530 ug/L

RT: 3.38 min Scan# 635

Delta R.T. -0.00 min

Lab File: VU039293.D

Acq: 06 Jul 2020 16:01




Tgt Ion: 73 Resp: 11472


Ion Ratio Lower Upper

73 100

43 26.6 17.5 26.3#

57 22.6 18.6 27.8





#18

trans-1,2-Dichloroethene

Concen: 0.141 ug/L

RT: 3.37 min Scan# 632

Delta R.T. -0.00 min

Lab File: VU039293.D

Acq: 06 Jul 2020 16:01

Instrument :

MSVOA_U

ClientSampleId :

H4279


Tot Ion: 96 Resp: 1008

Ion Ratio Lower Upper




96 100

61 131.6 104.7 194.5

98 50.4 43.1 80.1



Abundance Ion 96.00 (95.70 to 96.70): VU039293.D (-539) (-)



#19

1,1-Dichloroethane

Concen: 0.288 ug/L

RT: 3.90 min Scan# 795

Delta R.T. 0.01 min

Lab File: VU039293.D

Acq: 06 Jul 2020 16:01


Tgt Ion: 63 Resp: 4129

Ion Ratio Lower Upper



63 100


65 31.2 22.7 42.3

83 13.8 9.1 16.9



Abundance Ion 63.10 (62.80 to 63.80): VU039293.D (-700) (-)









#22
 cis-1,2-Dichloroethene
 Concen: 2.012 ug/L
 RT: 4.69 min Scan# 1041
 Delta R.T. -0.00 min
 Lab File: VU039293.D
 Acq: 06 Jul 2020 16:01

Instrument :
 MSVOA_U
 ClientSampleId :
 H4279


Tot Ion: 96 Resp: 16414
 Ion Ratio Lower Upper
 96 100
 61 141.8 93.7 173.9
 68 0.0 0.0 0.0



Abundance Ion 96.00 (95.70 to 96.70): VU0
 Ion 61.05 (60.75 to 61.75): VU0
 Ion 68.15 (67.85 to 68.85): VU0



#33
 Benzene
 Concen: 0.135 ug/L
 RT: 5.79 min Scan# 1384
 Delta R.T. -0.00 min
 Lab File: VU039293.D
 Acq: 06 Jul 2020 16:01
 Tgt Ion: 78 Resp: 4172



Abundance Ion 78.10 (77.80 to 78.80): VU0

