

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070718\
 Data File : VU025243.D
 Acq On : 06 Jul 2018 16:24
 Operator : MD/SY
 Sample : J3808-17DL 200X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021IW454-180628DL

Manual Integrations
 APPROVED

MMDadoda
 7/9/2018 12:57:59 PM

Quant Time: Jul 06 18:33:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	206955	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	313624	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	293909	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	155999	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	160218	47.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.84%	
35) Dibromofluoromethane	4.89	113	124036	47.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.30%	
50) Toluene-d8	7.57	98	452796	47.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.38%	
62) 4-Bromofluorobenzene	10.31	95	170474	45.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.44%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	481741	176.507	ug/l	98
6) Chloroethane	1.70	64	7684	4.070	ug/l	96
12) 1,1-Dichloroethene	2.29	96	2997	1.272	ug/l	96
21) trans-1,2-Dichloroethene	2.99	96	7527	2.900	ug/l	97
27) cis-1,2-Dichloroethene	4.23	96	1912082	640.807	ug/l	89
40) Benzene	5.40	78	12374	1.178	ug/l	97
44) Trichloroethene	6.19	130	20627	6.902	ug/l	96
65) Chlorobenzene	9.12	112	159803	21.284	ug/l	99
91) 1,2-Dichlorobenzene	11.88	146	58713	10.464	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070718\
Data File : VU025243.D
Acq On : 06 Jul 2018 16:24
Operator : MD/SY
Sample : J3808-17DL 200X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021IW454-180628DL

Manual Integrations
APPROVED

MMDadoda
7/9/2018 12:57:59 PM

Quant Time: Jul 06 18:33:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

