

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070718\
 Data File : VU025251.D
 Acq On : 06 Jul 2018 19:42
 Operator : MD/SY
 Sample : J3808-14
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WP021MW364-180628

Quant Time: Jul 07 00:22:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	197949	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	298121	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	276016	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	138576	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	150532	46.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.16%	
35) Dibromofluoromethane	4.89	113	116618	47.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.26%	
50) Toluene-d8	7.57	98	431661	47.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.66%	
62) 4-Bromofluorobenzene	10.31	95	156785	43.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.52%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.23	96	1731	0.607	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070718\
Data File : VU025251.D
Acq On : 06 Jul 2018 19:42
Operator : MD/SY
Sample : J3808-14
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WP021MW364-180628

Quant Time: Jul 07 00:22:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

