

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070718\
 Data File : VU025256.D
 Acq On : 06 Jul 2018 21:43
 Operator : MD/SY
 Sample : J3842-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 S10-TB-2018-3

Quant Time: Jul 07 00:39:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 13 13:55:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	195567	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	290155	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	271759	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	136352	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	147780	46.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.58%	
35) Dibromofluoromethane	4.89	113	113420	47.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.20%	
50) Toluene-d8	7.57	98	420071	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
62) 4-Bromofluorobenzene	10.31	95	153219	43.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.86%	
Target Compounds						
16) Acetone	2.32	43	1673	0.975	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070718\
Data File : VU025256.D
Acq On : 06 Jul 2018 21:43
Operator : MD/SY
Sample : J3842-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
S10-TB-2018-3

Quant Time: Jul 07 00:39:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U061318W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 13 13:55:26 2018
Response via : Initial Calibration

