

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070721\
 Data File : VU044324.D
 Acq On : 07 Jul 2021 19:55
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Jul 08 08:55:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U070721DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Jul 08 08:40:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 i	Fluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.346	0.357	-3.2	93	0.00
3 t	Chloromethane	0.334	0.336	-0.6	94	0.00
4 Rt	Vinyl Chloride	0.351	0.368	-4.8	94	0.00
5 T	Bromomethane	0.209	0.221	-5.7	98	0.00
6 T	Chloroethane	0.234	0.241	-3.0	95	0.00
7 T	Trichlorofluoromethane	0.531	0.558	-5.1	95	0.00
8	1,1,2-Trichloro-1,2,2-trifl	0.306	0.322	-5.2	94	0.00
9 Rt	1,1-Dichloroethene	0.267	0.280	-4.9	94	0.00
10 t	Iodomethane	0.282	0.328	-16.3	98	0.00
11 t	Allyl Chloride	0.475	0.475	0.0	92	0.00
12 t	Acrylonitrile	0.060	0.084	-40.0#	138	0.00
13 T	Acetone	0.040	0.062	-55.0#	153	0.00
14 T	Carbon Disulfide	0.723	0.748	-3.5	94	0.00
15 RT	Methylene Chloride	0.486	0.392	19.3	106	0.00
16 RT	trans-1,2-Dichloroethene	0.296	0.306	-3.4	95	0.00
17 t	1,1-Dichloroethane	0.597	0.633	-6.0	96	0.00
18 T	2-Butanone	0.078	0.116	-48.7#	147	0.00
19	Cyclohexane	0.478	0.521	-9.0	99	0.00
20	Methylcyclohexane	0.491	0.515	-4.9	94	0.00
21 T	2,2-Dichloropropane	0.582	0.551	5.3	88	0.00
22 RT	cis-1,2-Dichloroethene	0.349	0.362	-3.7	97	0.00
23 t	Diethyl Ether	0.223	0.239	-7.2	98	0.00
24 t	tert-Butyl Alcohol	0.017	0.032	-88.2#	195	0.01
25 t	Methyl tert-Butyl Ether	0.821	0.899	-9.5	98	0.00
26 t	Bromochloromethane	0.153	0.157	-2.6	94	0.00
27 t	Chloroform	0.632	0.649	-2.7	96	0.00
28 RT	1,1,1-Trichloroethane	0.558	0.589	-5.6	96	0.00
29 T	1,1-Dichloropropene	0.407	0.419	-2.9	93	0.00
30 RT	Carbon Tetrachloride	0.430	0.449	-4.4	93	0.00
31 t	Isopropyl Ether	1.023	1.085	-6.1	97	0.00
32	Ethyl-t-butyl ether	0.987	1.082	-9.6	96	0.00
33	Tert-Amyl methyl ether	0.796	0.835	-4.9	92	0.00
34 t	Propionitrile	0.020	0.033	-65.0#	159	0.00
35 RT	Benzene	1.126	1.178	-4.6	94	0.00
36 RT	1,2-Dichloroethane	0.368	0.383	-4.1	92	0.00
37 RT	Trichloroethene	0.314	0.328	-4.5	93	0.00
38 Rt	1,2-Dichloropropane	0.313	0.325	-3.8	94	0.00
39 t	Methacrylonitrile	0.123	0.141	-14.6	108	0.00
40 t	Methyl acrylate	0.171	0.227	-32.7#	124	0.00
41 t	Tetrahydrofuran	0.051	0.074	-45.1#	140	0.00
42 t	1-Chlorobutane	0.563	0.588	-4.4	96	0.00
43 T	Dibromomethane	0.165	0.171	-3.6	92	0.00
44 T	Bromodichloromethane	0.419	0.431	-2.9	92	0.00
45 T	4-Methyl-2-Pentanone	0.236	0.240	-1.7	92	0.00
46 t	t-1,4-Dichloro-2-butene	0.093	0.113	-21.5	101	0.00
47 t	Methyl methacrylate	0.169	0.174	-3.0	91	0.00
48 t	Ethyl methacrylate	0.362	0.369	-1.9	91	0.00
49 Rt	Toluene	0.740	0.780	-5.4	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070721\
 Data File : VU044324.D
 Acq On : 07 Jul 2021 19:55
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Jul 08 08:55:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U070721DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Jul 08 08:40:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 T	t-1,3-Dichloropropene	0.427	0.440	-3.0	90	0.00
51 T	cis-1,3-Dichloropropene	0.485	0.494	-1.9	90	0.00
52 RT	1,1,2-Trichloroethane	0.229	0.238	-3.9	92	0.00
53 t	1,3-Dichloropropane	0.421	0.436	-3.6	92	0.00
54 t	2-Hexanone	0.172	0.173	-0.6	91	0.00
55 t	Dibromochloromethane	0.295	0.312	-5.8	92	0.00
56 T	1,2-Dibromoethane	0.229	0.241	-5.2	93	0.00
57 S	4-Bromofluorobenzene	0.443	0.443	0.0	91	0.00
58 RT	Tetrachloroethene	0.290	0.309	-6.6	96	0.00
59 Rt	Chlorobenzene	0.837	0.881	-5.3	94	0.00
60 T	1,1,1,2-Tetrachloroethane	0.313	0.333	-6.4	96	0.00
61 t	Pentachloroethane	0.237	0.239	-0.8	87	0.00
62 t	Hexachloroethane	0.269	0.289	-7.4	93	0.00
63 Rt	Ethyl Benzene	1.481	1.562	-5.5	93	0.00
64 RT	m/p-Xylenes	0.575	0.608	-5.7	94	0.00
65 RT	o-Xylene	0.562	0.596	-6.0	94	0.00
66 RT	Styrene	0.965	1.022	-5.9	93	0.00
67 t	Bromoform	0.170	0.180	-5.9	90	0.00
68 S	1,2-Dichlorobenzene-d4	0.442	0.438	0.9	89	0.00
69 T	Isopropylbenzene	1.538	1.615	-5.0	93	0.00
70 T	1,1,2,2-Tetrachloroethane	0.308	0.317	-2.9	91	0.00
71 T	1,2,3-Trichloropropane	0.243	0.216	11.1	80	0.00
72 t	Bromobenzene	0.359	0.382	-6.4	93	0.00
73 t	n-propylbenzene	0.422	0.448	-6.2	93	0.00
74 t	2-Chlorotoluene	0.363	0.378	-4.1	93	0.00
75 t	1,3,5-Trimethylbenzene	1.316	1.390	-5.6	92	0.00
76 t	4-Chlorotoluene	0.382	0.399	-4.5	94	0.00
77 t	tert-Butylbenzene	1.323	1.385	-4.7	92	0.00
78 t	1,2,4-Trimethylbenzene	1.332	1.425	-7.0	93	0.00
79 t	sec-Butylbenzene	1.753	1.864	-6.3	93	0.00
80	Nitrobenzene	0.013	0.013	0.0	87	0.00
81 t	p-Isopropyltoluene	1.484	1.587	-6.9	93	0.00
82 t	1,3-Dichlorobenzene	0.731	0.765	-4.7	93	0.00
83 Rt	1,4-Dichlorobenzene	0.726	0.758	-4.4	92	0.00
84 t	n-Butylbenzene	1.421	1.546	-8.8	92	0.00
85 Rt	1,2-Dichlorobenzene	0.693	0.727	-4.9	93	0.00
86 t	1,2-Dibromo-3-Chloropropane	0.059	0.065	-10.2	95	0.00
87 Rt	1,2,4-Trichlorobenzene	0.497	0.534	-7.4	94	0.00
88 t	Hexachlorobutadiene	0.255	0.271	-6.3	93	0.00
89 t	Naphthalene	0.952	0.993	-4.3	91	0.00
90 t	1,2,3-Trichlorobenzene	0.439	0.481	-9.6	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0