

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU070722\  
 Data File : VU049655.D  
 Acq On : 07 Jul 2022 16:26  
 Operator : SY/MD  
 Sample : VSTDCCC010  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 LabSampleId :  
 VSTDCCC010

Quant Time: Jul 08 09:12:11 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\524U070622DW.M  
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER  
 QLast Update : Thu Jul 07 07:00:33 2022  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 30% Max. Rel. Area : 200%

|       | Compound                    | AvgRF | CCRF  | %Dev   | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|--------|-------|----------|
| 1 i   | Fluorobenzene               | 1.000 | 1.000 | 0.0    | 73    | 0.00     |
| 2 T   | Dichlorodifluoromethane     | 0.383 | 0.412 | -7.6   | 76    | 0.00     |
| 3 t   | Chloromethane               | 0.392 | 0.388 | 1.0    | 76    | 0.00     |
| 4 Rt  | Vinyl Chloride              | 0.357 | 0.375 | -5.0   | 76    | 0.00     |
| 5 T   | Bromomethane                | 0.186 | 0.164 | 11.8   | 68    | 0.00     |
| 6 T   | Chloroethane                | 0.200 | 0.204 | -2.0   | 74    | 0.00     |
| 7 T   | Trichlorofluoromethane      | 0.428 | 0.473 | -10.5  | 78    | 0.00     |
| 8     | 1,1,2-Trichloro-1,2,2-trifl | 0.225 | 0.237 | -5.3   | 74    | 0.00     |
| 9 Rt  | 1,1-Dichloroethene          | 0.227 | 0.226 | 0.4    | 70    | 0.00     |
| 10 t  | Iodomethane                 | 0.288 | 0.292 | -1.4   | 68    | 0.00     |
| 11 t  | Allyl Chloride              | 0.387 | 0.391 | -1.0   | 72    | 0.00     |
| 12 t  | Acrylonitrile               | 0.061 | 0.064 | -4.9   | 81    | 0.00     |
| 13 T  | Acetone                     | 0.048 | 0.049 | -2.1   | 83    | 0.00     |
| 14 T  | Carbon Disulfide            | 0.747 | 0.728 | 2.5    | 70    | 0.00     |
| 15 RT | Methylene Chloride          | 0.284 | 0.284 | 0.0    | 82    | 0.00     |
| 16 RT | trans-1,2-Dichloroethene    | 0.238 | 0.236 | 0.8    | 70    | 0.00     |
| 17 t  | 1,1-Dichloroethane          | 0.492 | 0.503 | -2.2   | 73    | 0.00     |
| 18 T  | 2-Butanone                  | 0.082 | 0.086 | -4.9   | 81    | 0.00     |
| 19    | Cyclohexane                 | 0.471 | 0.415 | 11.9   | 62    | 0.00     |
| 20    | Methylcyclohexane           | 0.464 | 0.463 | 0.2    | 70    | 0.00     |
| 21 T  | 2,2-Dichloropropane         | 0.403 | 0.406 | -0.7   | 70    | 0.00     |
| 22 RT | cis-1,2-Dichloroethene      | 0.294 | 0.292 | 0.7    | 70    | 0.00     |
| 23 t  | Diethyl Ether               | 0.193 | 0.199 | -3.1   | 75    | 0.00     |
| 24 t  | tert-Butyl Alcohol          | 0.022 | 0.029 | -31.8# | 99    | 0.00     |
| 25 t  | Methyl tert-Butyl Ether     | 0.634 | 0.665 | -4.9   | 77    | 0.00     |
| 26 t  | Bromochloromethane          | 0.124 | 0.130 | -4.8   | 76    | 0.00     |
| 27 t  | Chloroform                  | 0.485 | 0.510 | -5.2   | 76    | 0.00     |
| 28 RT | 1,1,1-Trichloroethane       | 0.427 | 0.456 | -6.8   | 75    | 0.00     |
| 29 T  | 1,1-Dichloropropene         | 0.382 | 0.387 | -1.3   | 72    | 0.00     |
| 30 RT | Carbon Tetrachloride        | 0.360 | 0.391 | -8.6   | 75    | 0.00     |
| 31 t  | Isopropyl Ether             | 0.844 | 0.850 | -0.7   | 71    | 0.00     |
| 32    | Ethyl-t-butyl ether         | 0.799 | 0.801 | -0.3   | 71    | 0.00     |
| 33    | Tert-Amyl methyl ether      | 0.695 | 0.708 | -1.9   | 74    | 0.00     |
| 34 t  | Propionitrile               | 0.023 | 0.027 | -17.4  | 84    | 0.00     |
| 35 RT | Benzene                     | 1.087 | 1.115 | -2.6   | 72    | 0.00     |
| 36 RT | 1,2-Dichloroethane          | 0.309 | 0.338 | -9.4   | 80    | 0.00     |
| 37 RT | Trichloroethene             | 0.275 | 0.269 | 2.2    | 69    | 0.00     |
| 38 Rt | 1,2-Dichloropropane         | 0.288 | 0.301 | -4.5   | 75    | 0.00     |
| 39 t  | Methacrylonitrile           | 0.100 | 0.106 | -6.0   | 77    | 0.00     |
| 40 t  | Methyl acrylate             | 0.157 | 0.168 | -7.0   | 76    | 0.00     |
| 41 t  | Tetrahydrofuran             | 0.059 | 0.059 | 0.0    | 83    | 0.00     |
| 42 t  | 1-Chlorobutane              | 0.535 | 0.555 | -3.7   | 73    | 0.00     |
| 43 T  | Dibromomethane              | 0.143 | 0.155 | -8.4   | 79    | 0.00     |
| 44 T  | Bromodichloromethane        | 0.344 | 0.374 | -8.7   | 76    | 0.00     |
| 45 T  | 4-Methyl-2-Pentanone        | 0.195 | 0.208 | -6.7   | 80    | 0.00     |
| 46 t  | t-1,4-Dichloro-2-butene     | 0.063 | 0.109 | -73.0# | 113   | 0.00     |
| 47 t  | Methyl methacrylate         | 0.145 | 0.154 | -6.2   | 76    | 0.00     |
| 48 t  | Ethyl methacrylate          | 0.290 | 0.300 | -3.4   | 74    | 0.00     |
| 49 Rt | Toluene                     | 0.664 | 0.672 | -1.2   | 71    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU070722\  
 Data File : VU049655.D  
 Acq On : 07 Jul 2022 16:26  
 Operator : SY/MD  
 Sample : VSTDCCC010  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 LabSampleId :  
 VSTDCCC010

Quant Time: Jul 08 09:12:11 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\524U070622DW.M  
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER  
 QLast Update : Thu Jul 07 07:00:33 2022  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 30% Max. Rel. Area : 200%

|       | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|-------|-------|----------|
| 50 T  | t-1,3-Dichloropropene       | 0.332 | 0.350 | -5.4  | 73    | 0.00     |
| 51 T  | cis-1,3-Dichloropropene     | 0.394 | 0.409 | -3.8  | 71    | 0.00     |
| 52 RT | 1,1,2-Trichloroethane       | 0.198 | 0.207 | -4.5  | 76    | 0.00     |
| 53 t  | 1,3-Dichloropropane         | 0.355 | 0.372 | -4.8  | 77    | 0.00     |
| 54 t  | 2-Hexanone                  | 0.135 | 0.146 | -8.1  | 80    | 0.00     |
| 55 t  | Dibromochloromethane        | 0.226 | 0.248 | -9.7  | 76    | 0.00     |
| 56 T  | 1,2-Dibromoethane           | 0.192 | 0.203 | -5.7  | 77    | 0.00     |
| 57 S  | 4-Bromofluorobenzene        | 0.366 | 0.352 | 3.8   | 71    | 0.00     |
| 58 RT | Tetrachloroethene           | 0.229 | 0.231 | -0.9  | 71    | 0.00     |
| 59 Rt | Chlorobenzene               | 0.695 | 0.697 | -0.3  | 71    | 0.00     |
| 60 T  | 1,1,1,2-Tetrachloroethane   | 0.247 | 0.264 | -6.9  | 75    | 0.00     |
| 61 t  | Pentachloroethane           | 0.190 | 0.201 | -5.8  | 74    | 0.00     |
| 62 t  | Hexachloroethane            | 0.190 | 0.198 | -4.2  | 71    | 0.00     |
| 63 Rt | Ethyl Benzene               | 1.250 | 1.294 | -3.5  | 71    | 0.00     |
| 64 RT | m/p-Xylenes                 | 0.463 | 0.478 | -3.2  | 71    | 0.00     |
| 65 RT | o-Xylene                    | 0.455 | 0.464 | -2.0  | 72    | 0.00     |
| 66 RT | Styrene                     | 0.730 | 0.778 | -6.6  | 72    | 0.00     |
| 67 t  | Bromoform                   | 0.129 | 0.141 | -9.3  | 77    | 0.00     |
| 68 S  | 1,2-Dichlorobenzene-d4      | 0.337 | 0.341 | -1.2  | 73    | 0.00     |
| 69 T  | Isopropylbenzene            | 1.214 | 1.246 | -2.6  | 70    | 0.00     |
| 70 T  | 1,1,2,2-Tetrachloroethane   | 0.257 | 0.269 | -4.7  | 79    | 0.00     |
| 71 T  | 1,2,3-Trichloropropane      | 0.213 | 0.223 | -4.7  | 86    | 0.00     |
| 72 t  | Bromobenzene                | 0.274 | 0.287 | -4.7  | 73    | 0.00     |
| 73 t  | n-propylbenzene             | 0.306 | 0.325 | -6.2  | 71    | 0.00     |
| 74 t  | 2-Chlorotoluene             | 0.273 | 0.283 | -3.7  | 72    | 0.00     |
| 75 t  | 1,3,5-Trimethylbenzene      | 1.015 | 1.052 | -3.6  | 71    | 0.00     |
| 76 t  | 4-Chlorotoluene             | 0.274 | 0.281 | -2.6  | 72    | 0.00     |
| 77 t  | tert-Butylbenzene           | 0.999 | 1.022 | -2.3  | 71    | 0.00     |
| 78 t  | 1,2,4-Trimethylbenzene      | 1.019 | 1.074 | -5.4  | 72    | 0.00     |
| 79 t  | sec-Butylbenzene            | 1.311 | 1.366 | -4.2  | 71    | 0.00     |
| 80    | Nitrobenzene                | 0.011 | 0.011 | 0.0   | 78    | 0.00     |
| 81 t  | p-Isopropyltoluene          | 1.063 | 1.123 | -5.6  | 71    | 0.00     |
| 82 t  | 1,3-Dichlorobenzene         | 0.529 | 0.548 | -3.6  | 72    | 0.00     |
| 83 Rt | 1,4-Dichlorobenzene         | 0.534 | 0.548 | -2.6  | 73    | 0.00     |
| 84 t  | n-Butylbenzene              | 0.990 | 1.078 | -8.9  | 71    | 0.00     |
| 85 Rt | 1,2-Dichlorobenzene         | 0.513 | 0.524 | -2.1  | 74    | 0.00     |
| 86 t  | 1,2-Dibromo-3-Chloropropane | 0.045 | 0.050 | -11.1 | 83    | 0.00     |
| 87 Rt | 1,2,4-Trichlorobenzene      | 0.346 | 0.372 | -7.5  | 71    | 0.00     |
| 88 t  | Hexachlorobutadiene         | 0.176 | 0.193 | -9.7  | 73    | 0.00     |
| 89 t  | Naphthalene                 | 0.725 | 0.759 | -4.7  | 72    | 0.00     |
| 90 t  | 1,2,3-Trichlorobenzene      | 0.330 | 0.351 | -6.4  | 72    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0