


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070722\
Data File : VU049644.D
Acq On : 07 Jul 2022 09:00
Operator : SY/MD
Sample : BFB
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U070622DW.M
Title : METHOD 524.2 VOLATILES DRINKING WATER
Last Update : Thu Jul 07 07:00:33 2022



AutoFind: Scans 2890, 2891, 2892; Background Corrected with Scan 2877

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.6	6738	PASS
75	95	30	60	53.8	19469	PASS
95	95	100	100	100.0	36216	PASS
96	95	5	9	7.0	2544	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	79.0	28608	PASS
175	174	5	9	7.5	2137	PASS
176	174	95	101	96.7	27653	PASS
177	176	5	9	6.6	1832	PASS