

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070820\
 Data File : VU039342.D
 Acq On : 08 Jul 2020 13:02
 Operator : SY/MD
 Sample : VU0708WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK29

Quant Time: Jul 09 02:38:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 09 02:29:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	114156	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	108123	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	52223	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	27076	4.09	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.80%
7) Chloroethane-d5	1.93	69	26091	3.85	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.00%
11) 1,1-Dichloroethene-d2	2.58	63	41067	3.16	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.20%
20) 2-Butanone-d5	4.66	46	135087	44.62	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.24%
24) Chloroform-d	5.09	84	58025	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.40%
26) 1,2-Dichloroethane-d4	5.72	65	38065	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.00%
32) Benzene-d6	5.75	84	116792	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.80%
36) 1,2-Dichloropropane-d6	6.70	67	39015	4.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.40%
41) Toluene-d8	7.91	98	102544	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.80%
43) trans-1,3-Dichloropropene-	8.19	79	16315	4.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.80%
46) 2-Hexanone-d5	8.64	63	97650	40.76	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.52%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	36234	4.00	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	41000	4.40	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070820\
Data File : VU039342.D
Acq On : 08 Jul 2020 13:02
Operator : SY/MD
Sample : VU0708WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK29

Quant Time: Jul 09 02:38:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 09 02:29:26 2020
Response via : Initial Calibration

