

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070820\
 Data File : VU039355.D
 Acq On : 08 Jul 2020 18:16
 Operator : SY/MD
 Sample : L3231-07
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H4293

Quant Time: Jul 09 04:58:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 09 02:29:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	104139	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	96791	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	47090	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	33135	5.49	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.80%
7) Chloroethane-d5	1.92	69	30653	4.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.20%
11) 1,1-Dichloroethene-d2	2.58	63	50136	4.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.40%
20) 2-Butanone-d5	4.67	46	139536	50.53	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.06%
24) Chloroform-d	5.09	84	72465	5.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.80%
26) 1,2-Dichloroethane-d4	5.72	65	44171	5.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.60%
32) Benzene-d6	5.74	84	139324	5.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	119.60%
36) 1,2-Dichloropropane-d6	6.71	67	48343	5.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	118.00%
41) Toluene-d8	7.91	98	120438	5.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.20%
43) trans-1,3-Dichloropropene-	8.19	79	18551	5.13	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.60%
46) 2-Hexanone-d5	8.65	63	107841	50.29	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.58%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	41728	5.14	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	48964	5.83	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.60%

Target Compounds

					Ovalue
5) Vinyl chloride	1.61	62	14081	1.658 ug/L	# 1
18) trans-1,2-Dichloroethene	3.37	96	6070	0.885 ug/L	93
22) cis-1,2-Dichloroethene	4.69	96	3570	0.464 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070820\
Data File : VU039355.D
Acq On : 08 Jul 2020 18:16
Operator : SY/MD
Sample : L3231-07
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4293

Quant Time: Jul 09 04:58:49 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 09 02:29:26 2020
Response via : Initial Calibration

