

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070822\
 Data File : VU049661.D
 Acq On : 08 Jul 2022 17:32
 Operator : SY/MD
 Sample : N3214-09 0.8ppb
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MDL-WATER-03-QT2-2022

Quant Time: Jul 09 04:56:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U070622DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Sat Jul 09 04:54:26 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Krupa Patel 07/11/2022
 Supervised By :Mahesh Dadoda 07/11/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.115	96	38065	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.632	95	12841	0.921	ug/l	0.00
Spiked Amount 1.000			Recovery	=	92.000%	
68) 1,2-Dichlorobenzene-d4	12.192	152	12842	1.000	ug/l	0.00
Spiked Amount 1.000			Recovery	=	100.000%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.385	85	10626	0.729	ug/l	100
3) Chloromethane	1.530	50	11062	0.742	ug/l	90
4) Vinyl Chloride	1.604	62	9736	0.716	ug/l	97
5) Bromomethane	1.858	94	5817	0.820	ug/l	93
6) Chloroethane	1.935	64	5467	0.718	ug/l	98
7) Trichlorofluoromethane	2.141	101	12371	0.759	ug/l	94
8) 1,1,2-Trichloro-1,2,2-...	2.578	101	5920	0.691	ug/l	97
9) 1,1-Dichloroethene	2.578	96	5893	0.683	ug/l	98
10) Iodomethane	2.723	142	4820	0.440	ug/l	90
11) Allyl Chloride	2.922	41	10811	0.735	ug/l	92
12) Acrylonitrile	3.327	53	3523m	1.527	ug/l	
13) Acetone	2.642	43	8390	4.608	ug/l	96
14) Carbon Disulfide	2.793	76	19246	0.677	ug/l	100
15) Methylene Chloride	3.044	84	11051	1.022	ug/l	94
16) trans-1,2-Dichloroethene	3.350	96	6264	0.691	ug/l	92
17) 1,1-Dichloroethane	3.867	63	13589	0.725	ug/l	95
18) 2-Butanone	4.735	43	12366	3.985	ug/l	99
19) Cyclohexane	5.378	56	11054m	0.617	ug/l	
20) Methylcyclohexane	6.761	83	9143	0.518	ug/l	91
21) 2,2-Dichloropropane	4.661	77	10217	0.666	ug/l	95
22) cis-1,2-Dichloroethene	4.668	96	7949	0.711	ug/l	98
23) Diethyl Ether	2.379	59	5525	0.751	ug/l	93
24) tert-Butyl Alcohol	3.237	59	7608	9.070	ug/l #	91
25) Methyl tert-Butyl Ether	3.369	73	18011	0.746	ug/l #	83
26) Bromochloromethane	4.977	128	3638	0.773	ug/l	96
27) Chloroform	5.092	83	13538	0.734	ug/l	96
28) 1,1,1-Trichloroethane	5.311	97	11879	0.730	ug/l	99
29) 1,1-Dichloropropene	5.523	75	9750	0.671	ug/l	97
30) Carbon Tetrachloride	5.520	117	9779	0.714	ug/l	96
31) Isopropyl Ether	4.005	45	22220	0.691	ug/l	92
32) Ethyl-t-butyl ether	4.510	59	20154	0.662	ug/l	96
33) Tert-Amyl methyl ether	5.948	73	16160	0.611	ug/l	99
34) Propionitrile	4.806	54	3112	3.497	ug/l #	93
35) Benzene	5.771	78	27462	0.664	ug/l	99
36) 1,2-Dichloroethane	5.796	62	9174	0.780	ug/l	98
37) Trichloroethene	6.539	130	7045	0.673	ug/l	92
38) 1,2-Dichloropropane	6.793	63	7872	0.718	ug/l	93
39) Methacrylonitrile	4.989	41	2628	0.689	ug/l #	93
40) Methyl acrylate	4.864	55	4287	0.718	ug/l #	81
41) Tetrahydrofuran	5.070	42	3508	1.556	ug/l #	87
42) 1-Chlorobutane	5.456	56	13450	0.660	ug/l	96
43) Dibromomethane	6.919	93	3834	0.702	ug/l	98
44) Bromodichloromethane	7.108	83	9171	0.701	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070822\
 Data File : VU049661.D
 Acq On : 08 Jul 2022 17:32
 Operator : SY/MD
 Sample : N3214-09 0.8ppb
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MDL-WATER-03-QT2-2022


Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 07/11/2022
 Supervised By :Mahesh Dadoda 07/11/2022

Quant Time: Jul 09 04:56:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U070622DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Sat Jul 09 04:54:26 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
45) 4-Methyl-2-Pentanone	7.803	43	25753	3.475	ug/l	95
46) t-1,4-Dichloro-2-butene	10.835	75	3233m	1.341	ug/l	
47) Methyl methacrylate	6.967	69	6832	1.241	ug/l	92
48) Ethyl methacrylate	8.340	69	6885	0.623	ug/l	98
49) Toluene	7.970	92	15984	0.632	ug/l	98
50) t-1,3-Dichloropropene	8.211	75	7977	0.631	ug/l	99
51) cis-1,3-Dichloropropene	7.610	75	9695	0.647	ug/l #	87
52) 1,1,2-Trichloroethane	8.404	97	5788	0.768	ug/l	92
53) 1,3-Dichloropropane	8.578	76	9718	0.719	ug/l	98
54) 2-Hexanone	8.693	43	17288	3.359	ug/l	94
55) Dibromochloromethane	8.812	129	6058	0.704	ug/l	92
56) 1,2-Dibromoethane	8.925	107	5061	0.694	ug/l	95
58) Tetrachloroethene	8.555	164	6082	0.699	ug/l	93
59) Chlorobenzene	9.449	112	16653	0.630	ug/l	98
60) 1,1,1,2-Tetrachloroethane	9.536	131	6635	0.705	ug/l	93
61) Pentachloroethane	11.426	117	5365	0.743	ug/l	97
62) Hexachloroethane	12.475	117	4663	0.645	ug/l	93
63) Ethyl Benzene	9.571	91	27977	0.588	ug/l	99
64) m/p-Xylenes	9.693	106	20971	1.191	ug/l	97
65) o-Xylene	10.102	106	10156	0.587	ug/l	96
66) Styrene	10.115	104	15925	0.573	ug/l	97
67) Bromoform	10.291	173	3267	0.665	ug/l	96
69) Isopropylbenzene	10.484	105	26204	0.567	ug/l	97
70) 1,1,2,2-Tetrachloroethane	10.783	83	7187	0.734	ug/l	92
71) 1,2,3-Trichloropropane	10.822	75	6277m	0.775	ug/l	
72) Bromobenzene	10.783	156	7314	0.702	ug/l	99
73) n-propylbenzene	10.906	120	6493	0.557	ug/l	89
74) 2-Chlorotoluene	10.989	126	6485	0.624	ug/l	97
75) 1,3,5-Trimethylbenzene	11.089	105	20853	0.540	ug/l	99
76) 4-Chlorotoluene	11.098	126	6567	0.630	ug/l	79
77) tert-Butylbenzene	11.420	119	21774	0.572	ug/l	98
78) 1,2,4-Trimethylbenzene	11.468	105	21184	0.546	ug/l	97
79) sec-Butylbenzene	11.642	105	28473	0.571	ug/l	99
80) Nitrobenzene	13.217	77	1914	4.779	ug/l #	61
81) p-Isopropyltoluene	11.793	119	21517	0.532	ug/l	97
82) 1,3-Dichlorobenzene	11.748	146	13267	0.659	ug/l	96
83) 1,4-Dichlorobenzene	11.838	146	13537m	0.666	ug/l	
84) n-Butylbenzene	12.208	91	20239	0.537	ug/l	99
85) 1,2-Dichlorobenzene	12.211	146	13688	0.701	ug/l	98
86) 1,2-Dibromo-3-Chloropr...	12.999	75	1268	0.743	ug/l #	81
87) 1,2,4-Trichlorobenzene	13.841	180	8668	0.658	ug/l	97
88) Hexachlorobutadiene	14.021	225	4431	0.660	ug/l	94
89) Naphthalene	14.092	128	18134	0.657	ug/l	97
90) 1,2,3-Trichlorobenzene	14.333	180	8712	0.693	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.115 min Scan# 14
Delta R.T. -0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022



Tgt Ion: 96 Resp: 3806
Ion Ratio Lower Upper
96 100
70 20.9 16.7 25.1
95 8.6 7.4 11.0
97 6.3 0.0 0.0#




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#2
Dichlorodifluoromethane
Concen: 0.729 ug/l
RT: 1.385 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32


Tgt Ion: 85 Resp: 10626
Ion Ratio Lower Upper
85 100
87 32.4 16.3 48.8





#3
Chloromethane
Concen: 0.742 ug/l
RT: 1.530 min Scan# 58
Delta R.T. 0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022



Tgt Ion: 50 Resp: 11062
Ion Ratio Lower Upper
50 100
52 38.1 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#4
Vinyl Chloride
Concen: 0.716 ug/l
RT: 1.604 min Scan# 82
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32


Tgt Ion: 62 Resp: 9736
Ion Ratio Lower Upper
62 100
64 30.3 25.5 38.3





#5
Bromomethane
Concen: 0.820 ug/l
RT: 1.858 min Scan# 161
Delta R.T. 0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022




Tgt Ion: 94 Resp: 581
Ion Ratio Lower Upper
94 100
96 85.0 73.0 109.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022




#6
Chloroethane
Concen: 0.718 ug/l
RT: 1.935 min Scan# 185
Delta R.T. 0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Tgt Ion: 64 Resp: 5467
Ion Ratio Lower Upper
64 100
66 31.0 25.8 38.6





#7
Trichlorofluoromethane
Concen: 0.759 ug/l
RT: 2.141 min Scan# 248
Delta R.T. 0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32



Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022



Tgt Ion:101 Resp: 1237
Ion Ratio Lower Upper
101 100
103 69.2 51.8 77.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#8
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 0.691 ug/l
RT: 2.578 min Scan# 385
Delta R.T. -0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32


Tgt Ion:101 Resp: 5920
Ion Ratio Lower Upper
101 100
85 46.9 36.4 54.6
151 76.4 59.1 88.7





#9
1,1-Dichloroethene
Concen: 0.683 ug/l
RT: 2.578 min Scan# 385
Delta R.T. -0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument : MSVOA_U
ClientSampleId : MDL-WATER-03-QT2-2022



Tgt Ion: 96 Resp: 589
Ion Ratio Lower Upper
96 100
61 172.1 0.0 510.6
98 59.5 0.0 124.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#10
Iodomethane
Concen: 0.440 ug/l
RT: 2.723 min Scan# 430
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32

Tgt Ion:142 Resp: 4820
Ion Ratio Lower Upper
142 100
127 54.3 38.2 57.2





#11

Allyl Chloride

Concen: 0.735 ug/l

RT: 2.922 min Scan# 492

Delta R.T. 0.000 min

Lab File: VU049661.D


Acq: 08 Jul 2022 17:32

Instrument :

MSVOA_U

ClientSampleId :

MDL-WATER-03-QT2-2022



Tgt Ion: 41 Resp: 1081

Ion Ratio Lower Upper

41 100




39 72.7 53.2 79.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 07/11/2022

Supervised By :Mahesh Dadoda 07/11/2022



#12

Acrylonitrile

Concen: 1.527 ug/l m

RT: 3.327 min Scan# 618

Delta R.T. 0.010 min

Lab File: VU049661.D

Acq: 08 Jul 2022 17:32




Tgt Ion: 53 Resp: 3523


Ion Ratio Lower Upper

53 100

52 84.2 64.6 97.0

51 19.9 30.7 46.1#





#13

Acetone

Concen: 4.608 ug/l

RT: 2.642 min Scan# 401

Delta R.T. 0.013 min

Lab File: VU049661.D


Acq: 08 Jul 2022 17:32

Instrument :

MSVOA_U

ClientSampleId :

MDL-WATER-03-QT2-2022



Tgt Ion: 43 Resp: 8390

Ion Ratio Lower Upper

43 100




58 30.4 26.2 39.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 07/11/2022

Supervised By :Mahesh Dadoda 07/11/2022



#14

Carbon Disulfide

Concen: 0.677 ug/l

RT: 2.793 min Scan# 452

Delta R.T. 0.003 min

Lab File: VU049661.D




Acq: 08 Jul 2022 17:32


Tgt Ion: 76 Resp: 19246

Ion Ratio Lower Upper

76 100


78 9.1 7.4 11.0





#15
Methylene Chloride
Concen: 1.022 ug/l
RT: 3.044 min Scan# 511
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022




Tgt Ion: 84 Resp: 1105
Ion Ratio Lower Upper
84 100
49 138.3 104.6 157.0
51 43.8 0.0 81.4
86 69.8 52.6 78.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022




#16
trans-1,2-Dichloroethene
Concen: 0.691 ug/l
RT: 3.350 min Scan# 625
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Tgt Ion: 96 Resp: 6264
Ion Ratio Lower Upper
96 100
61 166.5 123.4 185.0
98 68.2 51.6 77.4





#17
1,1-Dichloroethane
Concen: 0.725 ug/l
RT: 3.867 min Scan# 786
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022



Tgt Ion: 63 Resp: 13589
Ion Ratio Lower Upper
63 100
98 4.5 3.4 10.2
100 3.6 2.0 6.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#18
2-Butanone
Concen: 3.985 ug/l
RT: 4.735 min Scan# 1056
Delta R.T. 0.019 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32


Tgt Ion: 43 Resp: 12366
Ion Ratio Lower Upper
43 100
57 8.6 0.0 18.0





#19
Cyclohexane
Concen: 0.617 ug/l m
RT: 5.378 min Scan# 11054
Delta R.T. -0.006 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument : MSVOA_U
ClientSampleId : MDL-WATER-03-QT2-2022




Tgt Ion: 56 Resp: 11054
Ion Ratio Lower Upper
56 100
69 26.3 24.2 36.4
84 73.9 63.4 95.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022




#20
Methylcyclohexane
Concen: 0.518 ug/l
RT: 6.761 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32



Tgt Ion: 83 Resp: 9143
Ion Ratio Lower Upper
83 100
55 88.5 63.0 94.4
98 41.2 35.7 53.5





#21

2,2-Dichloropropane

Concen: 0.666 ug/l

RT: 4.661 min Scan# 1033

Delta R.T. -0.003 min

Lab File: VU049661.D

Acq: 08 Jul 2022 17:32

Instrument :

MSVOA_U

ClientSampleId :

MDL-WATER-03-QT2-2022

Tgt Ion: 77 Resp: 1021

Ion Ratio Lower Upper

77 100





97 20.4 18.5 27.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 07/11/2022

Supervised By :Mahesh Dadoda 07/11/2022



#22

cis-1,2-Dichloroethene

Concen: 0.711 ug/l

RT: 4.668 min Scan# 1035

Delta R.T. 0.000 min

Lab File: VU049661.D

Acq: 08 Jul 2022 17:32




Tgt Ion: 96 Resp: 7949


Ion Ratio Lower Upper

96 100

61 147.8 0.0 375.8


98 60.9 31.6 95.0





#23
Diethyl Ether
Concen: 0.751 ug/l
RT: 2.379 min Scan# 31
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022



Tgt Ion: 59 Resp: 552
Ion Ratio Lower Upper
59 100
45 79.9 56.0 84.0
74 69.8 57.6 86.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#24
tert-Butyl Alcohol
Concen: 9.070 ug/l
RT: 3.237 min Scan# 590
Delta R.T. 0.035 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32


Tgt Ion: 59 Resp: 7608
Ion Ratio Lower Upper
59 100
57 6.7 8.2 12.2#





#25
Methyl tert-Butyl Ether
Concen: 0.746 ug/l
RT: 3.369 min Scan# 61
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022



Tgt Ion: 73 Resp: 1801
Ion Ratio Lower Upper
73 100
43 27.7 15.9 23.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#26
Bromochloromethane
Concen: 0.773 ug/l
RT: 4.977 min Scan# 1131
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32

Tgt Ion:128 Resp: 3638
Ion Ratio Lower Upper
128 100
49 180.2 0.0 349.2
130 124.5 103.0 154.6





#27

Chloroform

Concen: 0.734 ug/l

RT: 5.092 min Scan# 1166

Delta R.T. 0.003 min

Lab File: VU049661.D

Acq: 08 Jul 2022 17:32

Instrument :

MSVOA_U

ClientSampleId :

MDL-WATER-03-QT2-2022

Tgt Ion: 83 Resp: 13538

Ion Ratio Lower Upper

83 100





85 63.9 0.0 133.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 07/11/2022

Supervised By :Mahesh Dadoda 07/11/2022



#28

1,1,1-Trichloroethane

Concen: 0.730 ug/l

RT: 5.311 min Scan# 1235

Delta R.T. -0.006 min

Lab File: VU049661.D

Acq: 08 Jul 2022 17:32




Tgt Ion: 97 Resp: 11879


Ion Ratio Lower Upper

97 100

99 63.2 32.1 96.5

61 43.5 21.8 65.4





#29

1,1-Dichloropropene

Concen: 0.671 ug/l

RT: 5.523 min Scan# 11

Delta R.T. -0.003 min

Lab File: VU049661.D

Acq: 08 Jul 2022 17:32

Instrument :

MSVOA_U

ClientSampleId :

MDL-WATER-03-QT2-2022

Tgt Ion: 75 Resp: 9750





Ion	Ratio	Lower	Upper
75	100		
110	34.7	17.3	52.0
77	27.0	24.6	37.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 07/11/2022

Supervised By :Mahesh Dadoda 07/11/2022



#30

Carbon Tetrachloride

Concen: 0.714 ug/l

RT: 5.520 min Scan# 1300




Delta R.T. -0.003 min


Lab File: VU049661.D

Acq: 08 Jul 2022 17:32

Tgt Ion:117 Resp: 9779


Ion	Ratio	Lower	Upper
117	100		
119	97.3	74.6	112.0
121	31.9	24.2	36.4





#31
Isopropyl Ether
Concen: 0.691 ug/l
RT: 4.005 min Scan# 827
Delta R.T. 0.006 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022



Tgt Ion: 45 Resp: 22220
Ion Ratio Lower Upper
45 100
43 52.7 23.8 71.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#32
Ethyl-t-butyl ether
Concen: 0.662 ug/l
RT: 4.510 min Scan# 986
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32


Tgt Ion: 59 Resp: 20154
Ion Ratio Lower Upper
59 100
87 38.3 32.5 48.7





#33
Tert-Amyl methyl ether
Concen: 0.611 ug/l
RT: 5.948 min Scan# 1432
Delta R.T. 0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument : MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022



Tgt Ion: 73 Resp: 16160
Ion Ratio Lower Upper
73 100
87 24.7 19.7 29.5
71 11.9 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#34
Propionitrile
Concen: 3.497 ug/l
RT: 4.806 min Scan# 1078
Delta R.T. 0.019 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32

Tgt Ion: 54 Resp: 3112
Ion Ratio Lower Upper
54 100
52 14.4 15.0 22.6#
55 13.8 12.6 19.0
40 4.1 4.2 6.4#









#35
Benzene
Concen: 0.664 ug/l
RT: 5.771 min Scan# 1379
Delta R.T. -0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32

Instrument : MSVOA_U
ClientSampleId : MDL-WATER-03-QT2-2022

Tgt Ion: 78 Resp: 27462
Ion Ratio Lower Upper
78 100
77 23.9 18.6 28.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#36
1,2-Dichloroethane
Concen: 0.780 ug/l
RT: 5.796 min Scan# 1386
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32


Tgt Ion: 62 Resp: 9174
Ion Ratio Lower Upper
62 100
98 8.6 7.4 11.0





#37
Trichloroethene
Concen: 0.673 ug/l
RT: 6.539 min Scan# 1617
Delta R.T. -0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022



Tgt Ion:130 Resp: 7049
Ion Ratio Lower Upper
130 100
95 96.7 83.8 125.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#38
1,2-Dichloropropane
Concen: 0.718 ug/l
RT: 6.793 min Scan# 1696
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32

Tgt Ion: 63 Resp: 7872
Ion Ratio Lower Upper
63 100
65 27.5 25.0 37.4





#39

Methacrylonitrile

Concen: 0.689 ug/l

RT: 4.989 min Scan# 1132

Delta R.T. 0.009 min

Lab File: VU049661.D


Acq: 08 Jul 2022 17:32

Instrument :

MSVOA_U

ClientSampleId :

MDL-WATER-03-QT2-2022






Tgt Ion:	41	Resp:	2628
Ion	Ratio	Lower	Upper
41	100		
67	76.0	59.4	89.0
39	58.4	42.2	63.2
52	19.1	23.7	35.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 07/11/2022

Supervised By :Mahesh Dadoda 07/11/2022



#40

Methyl acrylate

Concen: 0.718 ug/l




RT: 4.864 min Scan# 1096


Delta R.T. 0.013 min

Lab File: VU049661.D

Acq: 08 Jul 2022 17:32


Tgt Ion:	55	Resp:	4287
Ion	Ratio	Lower	Upper
55	100		
85	9.3	13.5	20.3
58	3.2	7.0	10.4
42	0.0	7.0	10.6





#41
Tetrahydrofuran
Concen: 1.556 ug/l
RT: 5.070 min Scan# 1158
Delta R.T. 0.007 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument : MSVOA_U
ClientSampleId : MDL-WATER-03-QT2-2022




Tgt Ion: 42 Resp: 3508
Ion Ratio Lower Upper
42 100
72 37.8 38.9 58.3
71 36.0 33.7 50.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022




#42
1-Chlorobutane
Concen: 0.660 ug/l
RT: 5.456 min Scan# 1280
Delta R.T. -0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Tgt Ion: 56 Resp: 13450
Ion Ratio Lower Upper
56 100
41 55.6 26.4 79.2





#43
Dibromomethane
Concen: 0.702 ug/l
RT: 6.919 min Scan# 1735
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022




Tgt Ion: 93 Resp: 3834
Ion Ratio Lower Upper
93 100
95 79.7 65.0 97.6
174 92.9 76.3 114.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022




#44
Bromodichloromethane
Concen: 0.701 ug/l
RT: 7.108 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32



Tgt Ion: 83 Resp: 9171
Ion Ratio Lower Upper
83 100
85 60.5 49.4 74.0
127 9.6 6.6 10.0





#45

4-Methyl-2-Pentanone

Concen: 3.475 ug/l

RT: 7.803 min Scan# 2008

Delta R.T. 0.006 min

Lab File: VU049661.D

Acq: 08 Jul 2022 17:32

Instrument :

MSVOA_U

ClientSampleId :





MDL-WATER-03-QT2-2022

Tgt Ion: 43	Resp: 2575
Ion Ratio	Lower Upper
43 100	
58 37.5	20.3 61.0
85 15.5	14.6 21.8

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 07/11/2022

Supervised By :Mahesh Dadoda 07/11/2022



#46

t-1,4-Dichloro-2-butene

Concen: 1.341 ug/l m




RT: 10.835 min Scan# 2953


Delta R.T. 0.000 min

Lab File: VU049661.D

Acq: 08 Jul 2022 17:32

Tgt Ion: 75	Resp: 3233
Ion Ratio	Lower Upper
75 100	
53 71.6	64.4 96.6
89 41.8	44.2 66.2#
88 35.5	32.8 49.2





#47

Methyl methacrylate

Concen: 1.241 ug/l

RT: 6.967 min Scan# 1750

Delta R.T. 0.003 min

Lab File: VU049661.D

Acq: 08 Jul 2022 17:32

Instrument :

MSVOA_U

ClientSampleId :

MDL-WATER-03-QT2-2022

Tgt Ion: 69 Resp: 6832

Ion Ratio Lower Upper

69 100

41 120.1 0.0 225.4

39 66.4 45.8 68.8





100 34.3 28.8 43.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 07/11/2022

Supervised By :Mahesh Dadoda 07/11/2022



#48

Ethyl methacrylate

Concen: 0.623 ug/l

RT: 8.340 min Scan# 2177

Delta R.T. 0.003 min

Lab File: VU049661.D




Acq: 08 Jul 2022 17:32


Tgt Ion: 69 Resp: 6885

Ion Ratio Lower Upper

69 100

41 63.3 32.3 96.9





#49

Toluene

Concen: 0.632 ug/l

RT: 7.970 min Scan# 2062

Delta R.T. 0.000 min

Lab File: VU049661.D


Acq: 08 Jul 2022 17:32

Instrument :

MSVOA_U

ClientSampleId :

MDL-WATER-03-QT2-2022



Tgt Ion: 92 Resp: 15984

Ion Ratio Lower Upper

92 100



91 175.4 138.6 207.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 07/11/2022

Supervised By :Mahesh Dadoda 07/11/2022



#50

t-1,3-Dichloropropene

Concen: 0.631 ug/l

RT: 8.211 min Scan# 2137

Delta R.T. 0.000 min

Lab File: VU049661.D




Acq: 08 Jul 2022 17:32


Tgt Ion: 75 Resp: 7977

Ion Ratio Lower Upper

75 100


77 31.5 24.7 37.1





#51
 cis-1,3-Dichloropropene
 Concen: 0.647 ug/l
 RT: 7.610 min Scan# 1950
 Delta R.T. 0.000 min
 Lab File: VU049661.D
 Acq: 08 Jul 2022 17:32




Instrument : MSVOA_U
 ClientSampleId : MDL-WATER-03-QT2-2022



Tgt Ion: 75 Resp: 969
 Ion Ratio Lower Upper
 75 100
 77 27.8 25.4 38.0
 39 55.1 35.1 52.7




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 07/11/2022
 Supervised By :Mahesh Dadoda 07/11/2022



#52
 1,1,2-Trichloroethane
 Concen: 0.768 ug/l
 RT: 8.404 min Scan# 2197
 Delta R.T. 0.003 min
 Lab File: VU049661.D
 Acq: 08 Jul 2022 17:32

Tgt Ion: 97 Resp: 5788
 Ion Ratio Lower Upper
 97 100
 83 78.1 70.6 106.0
 85 60.8 46.2 69.4
 99 56.7 50.0 75.0





#53

1,3-Dichloropropane

Concen: 0.719 ug/l

RT: 8.578 min Scan# 2251

Delta R.T. 0.000 min

Lab File: VU049661.D


Acq: 08 Jul 2022 17:32

Instrument :

MSVOA_U

ClientSampleId :

MDL-WATER-03-QT2-2022






Tgt Ion: 76	Resp: 9718
Ion Ratio	Lower Upper
76	100
78	31.4 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 07/11/2022

Supervised By :Mahesh Dadoda 07/11/2022



#54

2-Hexanone

Concen: 3.359 ug/l




RT: 8.693 min Scan# 2287


Delta R.T. 0.006 min

Lab File: VU049661.D

Acq: 08 Jul 2022 17:32


Tgt Ion: 43	Resp: 17288
Ion Ratio	Lower Upper
43	100
58	51.8 37.1 77.1
57	18.6 0.0 39.5





#55
Dibromochloromethane
Concen: 0.704 ug/l
RT: 8.812 min Scan# 21
Delta R.T. 0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022




Tgt Ion:129 Resp: 6058
Ion Ratio Lower Upper
129 100
127 83.1 61.2 91.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022




#56
1,2-Dibromoethane
Concen: 0.694 ug/l
RT: 8.925 min Scan# 2359
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32



Tgt Ion:107 Resp: 5061
Ion Ratio Lower Upper
107 100
109 98.7 0.0 187.2





#57

4-Bromofluorobenzene

Concen: 0.921 ug/l

RT: 10.632 min Scan# 2890

Delta R.T. 0.000 min

Lab File: VU049661.D


Acq: 08 Jul 2022 17:32

Instrument :

MSVOA_U

ClientSampleId :

MDL-WATER-03-QT2-2022





Tgt Ion: 95 Resp: 1284

Ion	Ratio	Lower	Upper
95	100		
174	74.8	62.4	93.6
176	72.7	56.2	84.2

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 07/11/2022

Supervised By :Mahesh Dadoda 07/11/2022



#58

Tetrachloroethene


Concen: 0.699 ug/l

RT: 8.555 min Scan# 2244

Delta R.T. 0.003 min



Lab File: VU049661.D


Acq: 08 Jul 2022 17:32



Tgt Ion:164 Resp: 6082


Ion	Ratio	Lower	Upper
164	100		
166	127.8	102.4	153.6
129	88.5	79.5	119.3
131	85.1	77.4	116.2





#59
Chlorobenzene
Concen: 0.630 ug/l
RT: 9.449 min Scan# 2521
Delta R.T. 0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument : MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022



Tgt Ion:112 Resp: 1665
Ion Ratio Lower Upper
112 100
114 32.2 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#60
1,1,1,2-Tetrachloroethane
Concen: 0.705 ug/l
RT: 9.536 min Scan# 2549
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32

Tgt Ion:131 Resp: 6635
Ion Ratio Lower Upper
131 100
133 91.0 78.3 117.5
119 61.9 53.2 79.8









#61
Pentachloroethane
Concen: 0.743 ug/l
RT: 11.426 min Scan# 3138
Delta R.T. -0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32

Instrument : MSVOA_U
ClientSampleId : MDL-WATER-03-QT2-2022

Tgt Ion:117 Resp: 536
Ion Ratio Lower Upper
117 100
167 80.1 66.2 99.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#62
Hexachloroethane
Concen: 0.645 ug/l
RT: 12.475 min Scan# 3463
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32


Tgt Ion:117 Resp: 4663
Ion Ratio Lower Upper
117 100
201 70.2 61.0 91.4





#63
Ethyl Benzene
Concen: 0.588 ug/l
RT: 9.571 min Scan# 2560
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument : MSVOA_U
ClientSampleId : MDL-WATER-03-QT2-2022




Tgt Ion: 91 Resp: 2797
Ion Ratio Lower Upper
91 100
106 30.1 24.4 36.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022




#64
m/p-Xylenes
Concen: 1.191 ug/l
RT: 9.693 min Scan# 2598
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32



Tgt Ion:106 Resp: 20971
Ion Ratio Lower Upper
106 100
91 205.2 167.8 251.6





#65

o-Xylene

Concen: 0.587 ug/l

RT: 10.102 min Scan# 2725

Delta R.T. 0.000 min

Lab File: VU049661.D

Acq: 08 Jul 2022 17:32

Instrument :

MSVOA_U

ClientSampleId :

MDL-WATER-03-QT2-2022

Tgt Ion:106 Resp: 10150

Ion Ratio Lower Upper

106 100





91 229.0 111.1 333.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 07/11/2022

Supervised By :Mahesh Dadoda 07/11/2022



#66

Styrene

Concen: 0.573 ug/l

RT: 10.115 min Scan# 2729

Delta R.T. 0.000 min

Lab File: VU049661.D

Acq: 08 Jul 2022 17:32




Tgt Ion:104 Resp: 15925


Ion Ratio Lower Upper

104 100

78 55.4 42.1 63.1


103 57.8 44.5 66.7





#67
Bromoform
Concen: 0.665 ug/l
RT: 10.291 min Scan# 21
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022




Tgt Ion:173 Resp: 326
Ion Ratio Lower Upper
173 100
175 50.4 38.4 57.6
254 10.8 7.7 11.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022




#68
1,2-Dichlorobenzene-d4
Concen: 1.000 ug/l
RT: 12.192 min Scan# 3375
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32



Tgt Ion:152 Resp: 12842
Ion Ratio Lower Upper
152 100
115 79.0 0.0 254.6
150 173.0 0.0 623.6









#69
Isopropylbenzene
Concen: 0.567 ug/l
RT: 10.484 min Scan# 21
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32

Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022

Tgt Ion:105 Resp: 26204
Ion Ratio Lower Upper
105 100
120 26.8 12.6 37.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#70
1,1,2,2-Tetrachloroethane
Concen: 0.734 ug/l
RT: 10.783 min Scan# 2937
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32


Tgt Ion: 83 Resp: 7187
Ion Ratio Lower Upper
83 100
131 11.4 7.7 11.5
85 71.0 51.2 76.8





#71
1,2,3-Trichloropropane
Concen: 0.775 ug/l m
RT: 10.822 min Scan# 2951
Delta R.T. -0.006 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022




Tgt Ion: 75 Resp: 627
Ion Ratio Lower Upper
75 100
77 0.0 0.0 112.4
110 34.6 0.0 70.6
97 16.7 0.0 37.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022




#72
Bromobenzene
Concen: 0.702 ug/l
RT: 10.783 min Scan# 2937
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32



Tgt Ion:156 Resp: 7314
Ion Ratio Lower Upper
156 100
77 172.4 0.0 344.4
158 98.9 0.0 192.2









#73
n-propylbenzene
Concen: 0.557 ug/l
RT: 10.906 min Scan# 2975
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32

Instrument : MSVOA_U
ClientSampleId : MDL-WATER-03-QT2-2022

Tgt Ion:120 Resp: 6491
Ion Ratio Lower Upper
120 100
91 489.0 367.8 551.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#74
2-Chlorotoluene
Concen: 0.624 ug/l
RT: 10.989 min Scan# 3001
Delta R.T. 0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32


Tgt Ion:126 Resp: 6485
Ion Ratio Lower Upper
126 100
91 316.4 0.0 620.6





#75
1,3,5-Trimethylbenzene
Concen: 0.540 ug/l
RT: 11.089 min Scan# 3032
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument : MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022



Tgt Ion:105 Resp: 2085
Ion Ratio Lower Upper
105 100
120 47.5 37.6 56.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#76
4-Chlorotoluene
Concen: 0.630 ug/l
RT: 11.098 min Scan# 3035
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32


Tgt Ion:126 Resp: 6567
Ion Ratio Lower Upper
126 100
91 323.1 0.0 738.8





#77
tert-Butylbenzene
Concen: 0.572 ug/l
RT: 11.420 min Scan# 3135
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022



Tgt Ion:119 Resp: 21774
Ion Ratio Lower Upper
119 100
91 62.9 30.4 91.3
134 23.0 17.8 26.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#78
1,2,4-Trimethylbenzene
Concen: 0.546 ug/l
RT: 11.468 min Scan# 3150
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32


Tgt Ion:105 Resp: 21184
Ion Ratio Lower Upper
105 100
120 45.1 21.6 65.0





#79
 sec-Butylbenzene
 Concen: 0.571 ug/l
 RT: 11.642 min Scan# 31
 Delta R.T. -0.003 min
 Lab File: VU049661.D
 Acq: 08 Jul 2022 17:32




Instrument : MSVOA_U
 ClientSampleId :
 MDL-WATER-03-QT2-2022




Tgt Ion:105 Resp: 2847
 Ion Ratio Lower Upper
 105 100
 134 19.9 15.4 23.0

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 07/11/2022
 Supervised By :Mahesh Dadoda 07/11/2022




#80
 Nitrobenzene
 Concen: 4.779 ug/l
 RT: 13.217 min Scan# 3694
 Delta R.T. 0.010 min
 Lab File: VU049661.D
 Acq: 08 Jul 2022 17:32




Tgt Ion: 77 Resp: 1914
 Ion Ratio Lower Upper
 77 100
 123 15.3 18.1 64.3#
 65 1.0 10.8 13.8#





#81
p-Isopropyltoluene
Concen: 0.532 ug/l
RT: 11.793 min Scan# 31
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022




Tgt Ion:119 Resp: 2151
Ion Ratio Lower Upper
119 100
134 26.6 20.5 30.7
91 23.2 19.8 29.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022




#82
1,3-Dichlorobenzene
Concen: 0.659 ug/l
RT: 11.748 min Scan# 3237
Delta R.T. 0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Tgt Ion:146 Resp: 13267
Ion Ratio Lower Upper
146 100
111 46.3 33.4 50.2
148 65.1 51.2 76.8





#83
1,4-Dichlorobenzene
Concen: 0.666 ug/l m
RT: 11.838 min Scan# 31
Delta R.T. 0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022




Tgt Ion:146 Resp: 1353
Ion Ratio Lower Upper
146 100
111 39.8 32.6 49.0
148 63.1 51.6 77.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022




#84
n-Butylbenzene
Concen: 0.537 ug/l
RT: 12.208 min Scan# 3380
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Tgt Ion: 91 Resp: 20239
Ion Ratio Lower Upper
91 100
92 53.5 42.3 63.5
134 25.2 19.6 29.4





#85
1,2-Dichlorobenzene
Concen: 0.701 ug/l
RT: 12.211 min Scan# 3381
Delta R.T. 0.000 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022



Tgt Ion:146 Resp: 1368
Ion Ratio Lower Upper
146 100
111 43.2 22.1 66.1
148 62.7 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#86
1,2-Dibromo-3-Chloropropane
Concen: 0.743 ug/l
RT: 12.999 min Scan# 3626
Delta R.T. 0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32

Tgt Ion: 75 Resp: 1268
Ion Ratio Lower Upper
75 100
155 63.6 66.8 100.2#
157 87.5 83.8 125.8





#87

1,2,4-Trichlorobenzene

Concen: 0.658 ug/l

RT: 13.841 min Scan# 3888

Delta R.T. 0.000 min

Lab File: VU049661.D

Acq: 08 Jul 2022 17:32

Instrument :

MSVOA_U

ClientSampleId :

MDL-WATER-03-QT2-2022





Tgt Ion:180	Resp:	866
Ion Ratio	Lower	Upper
180	100	
182	93.0	77.0 115.6
145	30.6	25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 07/11/2022

Supervised By :Mahesh Dadoda 07/11/2022



#88

Hexachlorobutadiene

Concen: 0.660 ug/l




RT: 14.021 min Scan# 3944


Delta R.T. 0.000 min

Lab File: VU049661.D

Acq: 08 Jul 2022 17:32


Tgt Ion:225	Resp:	4431
Ion Ratio	Lower	Upper
225	100	
223	58.9	50.6 76.0
227	68.0	50.7 76.1





#89
Naphthalene
Concen: 0.657 ug/l
RT: 14.092 min Scan# 3966
Delta R.T. 0.006 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32




Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-03-QT2-2022



Tgt Ion:128 Resp: 18134
Ion Ratio Lower Upper
128 100
127 14.2 10.1 15.1
129 11.0 8.6 12.8

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 07/11/2022
Supervised By :Mahesh Dadoda 07/11/2022



#90
1,2,3-Trichlorobenzene
Concen: 0.693 ug/l
RT: 14.333 min Scan# 4041
Delta R.T. 0.003 min
Lab File: VU049661.D
Acq: 08 Jul 2022 17:32

Tgt Ion:180 Resp: 8712
Ion Ratio Lower Upper
180 100
182 93.0 76.6 114.8
145 32.5 27.3 40.9

