



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070822\
 Data File : VU049656.D
 Acq On : 08 Jul 2022 12:26
 Operator : SY/MD
 Sample : BFB
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U070622DW.M
 Title : METHOD 524.2 VOLATILES DRINKING WATER
 Last Update : Thu Jul 07 07:00:33 2022



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2876

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	13240	PASS
75	95	30	60	53.0	36837	PASS
95	95	100	100	100.0	69453	PASS
96	95	5	9	6.2	4336	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	76.0	52781	PASS
175	174	5	9	7.6	3992	PASS
176	174	95	101	95.8	50589	PASS
177	176	5	9	6.6	3325	PASS