

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070824\
 Data File : VU059883.D
 Acq On : 08 Jul 2024 15:10
 Operator : MD/SY
 Sample : P3138-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Jul 09 02:37:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jul 09 02:28:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	162124	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	165714	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	73386	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	46303	4.107	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.200%
7) Chloroethane-d5	1.907	69	42177	3.888	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.800%
11) 1,1-Dichloroethene-d2	2.560	65	18209	3.893	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.800%
20) 2-Butanone-d5	4.618	46	131385	56.579	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.160%
24) Chloroform-d	5.055	84	99682	4.308	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.200%
26) 1,2-Dichloroethane-d4	5.695	65	51590	4.628	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.600%
32) Benzene-d6	5.721	84	198040	4.342	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
36) 1,2-Dichloropropane-d6	6.685	67	64566	4.620	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.400%
41) Toluene-d8	7.894	98	165600	3.906	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.200%
43) trans-1,3-Dichloroprop...	8.177	79	21585	4.932	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.600%
46) 2-Hexanone-d5	8.627	63	104625	57.541	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	115.080%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	58153	5.036	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	66548	5.095	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU070824\
Data File : VU059883.D
Acq On : 08 Jul 2024 15:10
Operator : MD/SY
Sample : P3138-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Jul 09 02:37:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jul 09 02:28:04 2024
Response via : Initial Calibration

