

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070920\
 Data File : VU039389.D
 Acq On : 09 Jul 2020 11:55
 Operator : SY/MD
 Sample : L3243-21
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Jul 10 09:42:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 10 00:19:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.27	114	101211	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	93112	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	44324	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	30997	5.29	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.80%
7) Chloroethane-d5	1.93	69	26797	4.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.40%
11) 1,1-Dichloroethene-d2	2.58	63	42896	3.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.40%
20) 2-Butanone-d5	4.66	46	132822	49.49	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.98%
24) Chloroform-d	5.09	84	57035	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
26) 1,2-Dichloroethane-d4	5.72	65	37631	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.00%
32) Benzene-d6	5.75	84	116833	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.20%
36) 1,2-Dichloropropane-d6	6.71	67	38444	4.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.60%
41) Toluene-d8	7.91	98	100476	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.40%
43) trans-1,3-Dichloropropene-	8.19	79	17157	4.93	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.60%
46) 2-Hexanone-d5	8.64	63	94624	45.87	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.74%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	35771	4.58	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	41690	5.27	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.40%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	1360	0.180 ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU070920\
Data File : VU039389.D
Acq On : 09 Jul 2020 11:55
Operator : SY/MD
Sample : L3243-21
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Jul 10 09:42:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 10 00:19:08 2020
Response via : Initial Calibration

