

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070920\
 Data File : VU039402.D
 Acq On : 09 Jul 2020 17:57
 Operator : SY/MD
 Sample : L3231-01DL 10X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H4283DL

Quant Time: Jul 10 01:31:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jul 10 00:19:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.27	114	95310	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	88365	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	44066	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	25162	4.56	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.20%
7) Chloroethane-d5	1.93	69	30934	5.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.40%
11) 1,1-Dichloroethene-d2	2.58	63	40808	3.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.20%
20) 2-Butanone-d5	4.65	46	152886	60.49	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.98%
24) Chloroform-d	5.09	84	64782	5.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.20%
26) 1,2-Dichloroethane-d4	5.72	65	42875	5.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	116.20%
32) Benzene-d6	5.75	84	126183	5.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	118.60%
36) 1,2-Dichloropropane-d6	6.71	67	44277	5.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	118.40%
41) Toluene-d8	7.91	98	108050	5.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.20%
43) trans-1,3-Dichloropropene-	8.19	79	17906	5.43	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	108.60%
46) 2-Hexanone-d5	8.64	63	109902	56.14	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	112.28%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	40297	5.44	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	45502	5.78	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.60%

Target Compounds

					Ovalue
5) Vinyl chloride	1.61	62	6652	0.856 ug/L	# 46
17) Methyl tert-butyl Ether	3.39	73	12265	0.687 ug/L	98
18) trans-1,2-Dichloroethene	3.37	96	4797	0.764 ug/L	90
22) cis-1,2-Dichloroethene	4.69	96	56504	8.030 ug/L	99
34) Trichloroethene	6.56	95	31124	4.557 ug/L	96
47) Tetrachloroethene	8.56	164	627	0.132 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU070920\
Data File : VU039402.D
Acq On : 09 Jul 2020 17:57
Operator : SY/MD
Sample : L3231-01DL 10X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4283DL

Quant Time: Jul 10 01:31:36 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 10 00:19:08 2020
Response via : Initial Calibration

