

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071018\
 Data File : VU025284.D
 Acq On : 10 Jul 2018 13:13
 Operator : MD/SY
 Sample : J3898-15
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CPTB02-180703

Quant Time: Jul 11 09:15:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U070918W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 10 09:34:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	212577	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	321530	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	298688	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	156556	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	161196	48.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.62%	
35) Dibromofluoromethane	4.89	113	126795	46.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.50%	
50) Toluene-d8	7.57	98	460055	49.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.06%	
62) 4-Bromofluorobenzene	10.31	95	170343	45.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.10%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071018\
Data File : VU025284.D
Acq On : 10 Jul 2018 13:13
Operator : MD/SY
Sample : J3898-15
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CPTB02-180703

Quant Time: Jul 11 09:15:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

