

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071018\
 Data File : VU025285.D
 Acq On : 10 Jul 2018 13:38
 Operator : MD/SY
 Sample : J3905-10
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jul 11 09:16:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U070918W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 10 09:34:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	205261	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	310473	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	286288	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	147786	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	154648	48.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.98%	
35) Dibromofluoromethane	4.89	113	121392	46.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.70%	
50) Toluene-d8	7.57	98	440614	49.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.26%	
62) 4-Bromofluorobenzene	10.31	95	161731	44.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.58%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071018\
Data File : VU025285.D
Acq On : 10 Jul 2018 13:38
Operator : MD/SY
Sample : J3905-10
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TRIP-BLANK

Quant Time: Jul 11 09:16:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

