

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071018\
 Data File : VU025293.D
 Acq On : 10 Jul 2018 16:53
 Operator : MD/SY
 Sample : J3898-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SS038MW077-180703

Quant Time: Jul 11 09:21:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U070918W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 10 09:34:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	202890	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	300028	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	282725	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	145583	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	154062	48.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.76%	
35) Dibromofluoromethane	4.89	113	116674	46.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.20%	
50) Toluene-d8	7.57	98	433862	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
62) 4-Bromofluorobenzene	10.31	95	158329	45.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.74%	
Target Compounds						
11) Tert butyl alcohol	2.82	59	7921	11.52	ug/l	# 94
19) Methyl tert-butyl Ether	3.00	73	10152	1.30	ug/l	92
65) Chlorobenzene	9.12	112	4994	0.71	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071018\
Data File : VU025293.D
Acq On : 10 Jul 2018 16:53
Operator : MD/SY
Sample : J3898-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SS038MW077-180703

Quant Time: Jul 11 09:21:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

