

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071020\
 Data File : VU039432.D
 Acq On : 10 Jul 2020 16:06
 Operator : SY/MD
 Sample : L3261-05DL 40X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BE504DL

Quant Time: Jul 11 03:34:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jul 11 02:49:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.27	114	94245	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	89935	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	42652	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	23156	4.24	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.80%
7) Chloroethane-d5	1.92	69	28163	5.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.80%
11) 1,1-Dichloroethene-d2	2.58	63	36297	3.38	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.60%
20) 2-Butanone-d5	4.65	46	147001	58.82	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.64%
24) Chloroform-d	5.09	84	59389	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.20%
26) 1,2-Dichloroethane-d4	5.72	65	39401	5.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.00%
32) Benzene-d6	5.74	84	114819	5.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.00%
36) 1,2-Dichloropropane-d6	6.71	67	39490	5.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.80%
41) Toluene-d8	7.91	98	95857	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.20%
43) trans-1,3-Dichloropropene-	8.19	79	15827	4.71	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.20%
46) 2-Hexanone-d5	8.64	63	106432	53.41	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.82%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	38968	5.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	40568	5.33	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.60%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
8) Chloroethane	1.94	64	2006	0.361	ug/L #	80
22) cis-1,2-Dichloroethene	4.69	96	2509	0.361	ug/L	100
33) Benzene	5.79	78	169473	6.303	ug/L	100
42) Toluene	7.98	91	255631	8.878	ug/L	99
51) Chlorobenzene	9.45	112	21841	1.241	ug/L	96
52) Ethylbenzene	9.57	91	36489	1.147	ug/L	98
53) m,p-Xylene	9.70	106	36893	3.077	ug/L	100
54) o-Xylene	10.10	106	11665	0.999	ug/L	94
56) Isopropylbenzene	10.48	105	6441	0.204	ug/L	97
63) 1,4-Dichlorobenzene	11.83	146	1431	0.107	ug/L	95
65) 1,2-Dichlorobenzene	12.20	146	1438	0.107	ug/L #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071020\
Data File : VU039432.D
Acq On : 10 Jul 2020 16:06
Operator : SY/MD
Sample : L3261-05DL 40X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BE504DL

Quant Time: Jul 11 03:34:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jul 11 02:49:36 2020
Response via : Initial Calibration

