

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071020\
 Data File : VU039454.D
 Acq On : 11 Jul 2020 01:37
 Operator : SY/MD
 Sample : L3268-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F4A86

Quant Time: Jul 11 05:15:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jul 11 05:00:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.27	114	91439	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	84601	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	40253	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	19537	3.69	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.80%
7) Chloroethane-d5	1.92	69	24629	4.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.80%
11) 1,1-Dichloroethene-d2	2.58	63	31743	3.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.80%
20) 2-Butanone-d5	4.66	46	133087	54.88	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.76%
24) Chloroform-d	5.08	84	55022	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
26) 1,2-Dichloroethane-d4	5.72	65	36879	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.20%
32) Benzene-d6	5.75	84	100434	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.60%
36) 1,2-Dichloropropane-d6	6.71	67	36940	5.16	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.20%
41) Toluene-d8	7.91	98	87042	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
43) trans-1,3-Dichloropropene-	8.19	79	13756	4.35	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.00%
46) 2-Hexanone-d5	8.64	63	96535	51.50	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.00%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	35474	5.00	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	35562	4.95	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.00%

Target Compounds					Ovalue
13) Acetone	2.66	43	8971	6.854 ug/L	68
25) Chloroform	5.11	83	7901	0.664 ug/L	89
30) Cyclohexane	5.41	56	1853	0.177 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071020\
Data File : VU039454.D
Acq On : 11 Jul 2020 01:37
Operator : SY/MD
Sample : L3268-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F4A86

Quant Time: Jul 11 05:15:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR070720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jul 11 05:00:40 2020
Response via : Initial Calibration

