

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071024\
 Data File : VU059892.D
 Acq On : 10 Jul 2024 10:58
 Operator : MD/SY
 Sample : P3137-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Jul 11 01:53:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jul 11 01:53:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	140547	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	147200	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	63160	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	49397	5.055	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	101.000%	
7) Chloroethane-d5	1.907	69	41264	4.388	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	87.800%	
11) 1,1-Dichloroethene-d2	2.560	65	18234	4.497	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	90.000%	
20) 2-Butanone-d5	4.631	46	94910	47.146	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	94.300%	
24) Chloroform-d	5.055	84	84536	4.214	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	84.200%	
26) 1,2-Dichloroethane-d4	5.695	65	40909	4.234	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.600%	
32) Benzene-d6	5.721	84	171201	4.226	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	84.600%	
36) 1,2-Dichloropropane-d6	6.685	67	54815	4.416	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	88.400%	
41) Toluene-d8	7.894	98	142847	3.793	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	75.800%	
43) trans-1,3-Dichloroprop...	8.177	79	17791	4.577	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	91.600%	
46) 2-Hexanone-d5	8.631	63	72257	44.738	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	89.480%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	44238	4.313	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	86.200%	
66) 1,2-Dichlorobenzene-d4	12.190	152	53366	4.747	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	95.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071024\
Data File : VU059892.D
Acq On : 10 Jul 2024 10:58
Operator : MD/SY
Sample : P3137-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Jul 11 01:53:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR061724WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jul 11 01:53:22 2024
Response via : Initial Calibration

