

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071025\
 Data File : VU063469.D
 Acq On : 10 Jul 2025 10:18
 Operator : MD/SY
 Sample : VU0710WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0710WBL01

Quant Time: Jul 11 04:45:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 11 04:42:35 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.361	168	125158	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.242	114	219712	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.409	117	204489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.807	152	101767	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.692	65	77729	48.676	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.360%
35) Dibromofluoromethane	5.280	113	62022	46.831	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.660%
50) Toluene-d8	7.891	98	200240	46.383	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.760%
62) 4-Bromofluorobenzene	10.627	95	86116	45.605	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071025\
Data File : VU063469.D
Acq On : 10 Jul 2025 10:18
Operator : MD/SY
Sample : VU0710WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0710WBL01

Quant Time: Jul 11 04:45:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070825W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 11 04:42:35 2025
Response via : Initial Calibration

