

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071025\
 Data File : VU063472.D
 Acq On : 10 Jul 2025 11:27
 Operator : MD/SY
 Sample : VU0710WBL04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0710WBL04

Quant Time: Jul 11 04:47:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 11 04:42:35 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.357	168	116471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.238	114	203764	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.409	117	185746	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.804	152	96537	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.692	65	75847	51.040	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.080%
35) Dibromofluoromethane	5.274	113	60236	49.042	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.080%
50) Toluene-d8	7.888	98	192081	47.976	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.960%
62) 4-Bromofluorobenzene	10.624	95	85975	49.093	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071025\
Data File : VU063472.D
Acq On : 10 Jul 2025 11:27
Operator : MD/SY
Sample : VU0710WBL04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0710WBL04

Quant Time: Jul 11 04:47:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070825W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 11 04:42:35 2025
Response via : Initial Calibration

