

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071118\
 Data File : VU025316.D
 Acq On : 11 Jul 2018 15:51
 Operator : MD/SY
 Sample : J3922-13
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CPTB04-180709

Quant Time: Jul 12 09:08:35 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 10 09:34:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	200555	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	306554	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	285138	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	143958	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	152793	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
35) Dibromofluoromethane	4.89	113	119765	46.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.62%	
50) Toluene-d8	7.57	98	437177	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
62) 4-Bromofluorobenzene	10.31	95	160025	44.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.76%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071118\
Data File : VU025316.D
Acq On : 11 Jul 2018 15:51
Operator : MD/SY
Sample : J3922-13
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CPTB04-180709

Quant Time: Jul 12 09:08:35 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

