

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU071118\
 Data File : VU025317.D
 Acq On : 11 Jul 2018 16:16
 Operator : MD/SY
 Sample : J3936-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB-01-180710

Quant Time: Jul 12 09:09:04 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 10 09:34:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	208293	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	308370	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	290093	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	147955	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	155363	48.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.02%	
35) Dibromofluoromethane	4.89	113	122661	47.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.30%	
50) Toluene-d8	7.57	98	446207	50.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.18%	
62) 4-Bromofluorobenzene	10.31	95	163048	45.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.92%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU071118\
Data File : VU025317.D
Acq On : 11 Jul 2018 16:16
Operator : MD/SY
Sample : J3936-01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB-01-180710

Quant Time: Jul 12 09:09:04 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

