

Data File : VU025318.D
Acq On : 11 Jul 2018 16:40
Operator : MD/SY
Sample : J3936-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
OUTFALL001-180619

Quant Time: Jul 12 08:05:16 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	209698	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	311462	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	287925	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	149785	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	155195	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
35) Dibromofluoromethane	4.89	113	121951	46.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.82%	
50) Toluene-d8	7.57	98	449205	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
62) 4-Bromofluorobenzene	10.31	95	165440	45.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.34%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.23	96	6792	2.417	ug/l	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU025318.D
Acq On : 11 Jul 2018 16:40
Operator : MD/SY
Sample : J3936-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
OUTFALL001-180619

Quant Time: Jul 12 08:05:16 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

