

Data File : VU025319.D
Acq On : 11 Jul 2018 17:05
Operator : MD/SY
Sample : J3921-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
KY038PS027-180709

Quant Time: Jul 12 08:18:09 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	207661	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	309249	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	287602	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	146826	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	159512	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
35) Dibromofluoromethane	4.89	113	122168	46.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.66%	
50) Toluene-d8	7.57	98	448668	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
62) 4-Bromofluorobenzene	10.31	95	163752	45.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.06%	
Target Compounds						
16) Acetone	2.32	43	33411	24.767	ug/l	100
25) 2-Butanone	4.27	43	220160	98.453	ug/l	99
27) cis-1,2-Dichloroethene	4.23	96	112973	40.590	ug/l	91
29) Tetrahydrofuran	4.64	42	48891	36.450	ug/l	99
44) Trichloroethene	6.19	130	20685	7.271	ug/l	98
64) Tetrachloroethene	8.23	164	31155	11.884	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU025319.D
Acq On : 11 Jul 2018 17:05
Operator : MD/SY
Sample : J3921-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
KY038PS027-180709

Quant Time: Jul 12 08:18:09 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

