

Data File : VU025322.D
Acq On : 11 Jul 2018 18:18
Operator : MD/SY
Sample : J3921-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
KY038PS036-180709

Quant Time: Jul 12 04:46:01 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	200429	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	300483	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	277067	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	143190	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	151499	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
35) Dibromofluoromethane	4.89	113	118533	46.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.52%	
50) Toluene-d8	7.57	98	429943	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
62) 4-Bromofluorobenzene	10.31	95	158120	45.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.50%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.24	96	5015	1.867	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU025322.D
Acq On : 11 Jul 2018 18:18
Operator : MD/SY
Sample : J3921-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
KY038PS036-180709

Quant Time: Jul 12 04:46:01 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

