

Data File : VU025323.D
Acq On : 11 Jul 2018 18:43
Operator : MD/SY
Sample : J3922-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SS037MW202-180709

Quant Time: Jul 12 08:26:17 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 10 09:34:16 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	197914	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	301875	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	273675	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	140693	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	151613	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
35) Dibromofluoromethane	4.89	113	117067	45.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.94%	
50) Toluene-d8	7.57	98	430528	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
62) 4-Bromofluorobenzene	10.31	95	156043	44.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.90%	
Target Compounds						
4) Vinyl Chloride	1.40	62	12555	5.578	ug/l	98
27) cis-1,2-Dichloroethene	4.24	96	11977	4.515	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU025323.D

Acq On : 11 Jul 2018 18:43

Operator : MD/SY

Sample : J3922-09

Misc : 5.0mL/MSVOA_U/WATER

ALS Vial : 19 Sample Multiplier: 1

Instrument :

MSVOA_U

ClientSampleId :

SS037MW202-180709

Quant Time: Jul 12 08:26:17 2018

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U070918W.M

Quant Title : SW846 8260

QLast Update : Tue Jul 10 09:34:16 2018

Response via : Initial Calibration

