

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071122\
 Data File : VU049672.D
 Acq On : 11 Jul 2022 13:30
 Operator : SY/MD
 Sample : VU0711MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK148

Quant Time: Jul 12 02:33:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jul 12 02:23:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	230582	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	210679	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	91395	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	94661	48.656	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.320%
7) Chloroethane-d5	1.906	69	69896	49.702	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.400%
11) 1,1-Dichloroethene-d2	2.565	63	124999	36.374	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.740%
21) 2-Butanone-d5	4.620	46	170118	102.429	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.430%
24) Chloroform-d	5.063	84	164236	47.662	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.320%
26) 1,2-Dichloroethane-d4	5.700	65	112888	48.405	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.820%
32) Benzene-d6	5.726	84	323874	50.602	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.200%
36) 1,2-Dichloropropane-d6	6.690	67	111388	51.741	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.480%
41) Toluene-d8	7.896	98	276216	49.557	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.120%
43) trans-1,3-Dichloroprop...	8.179	79	45158	49.787	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.580%
47) 2-Hexanone-d5	8.632	63	98021	101.221	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.220%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	148546	49.314	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.620%
66) 1,2-Dichlorobenzene-d4	12.192	152	89401	51.881	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071122\
Data File : VU049672.D
Acq On : 11 Jul 2022 13:30
Operator : SY/MD
Sample : VU0711MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK148

Quant Time: Jul 12 02:33:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jul 12 02:23:25 2022
Response via : Initial Calibration

