

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071122\
 Data File : VU049675.D
 Acq On : 11 Jul 2022 14:58
 Operator : SY/MD
 Sample : N3639-01 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COL62

Quant Time: Jul 12 02:34:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jul 12 02:23:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	229444	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	208786	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	93063	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	90154	46.569	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.140%
7) Chloroethane-d5	1.906	69	66421	47.465	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.940%
11) 1,1-Dichloroethene-d2	2.562	63	118482	34.648	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.300%
21) 2-Butanone-d5	4.620	46	168703	102.081	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.080%
24) Chloroform-d	5.063	84	159870	46.625	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.260%
26) 1,2-Dichloroethane-d4	5.700	65	108859	46.909	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.820%
32) Benzene-d6	5.726	84	314227	49.540	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.080%
36) 1,2-Dichloropropane-d6	6.690	67	106702	50.013	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.020%
41) Toluene-d8	7.896	98	264815	47.942	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.880%
43) trans-1,3-Dichloroprop...	8.179	79	43233	48.097	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.200%
47) 2-Hexanone-d5	8.632	63	97462	101.556	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.560%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	147943	49.559	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.120%
66) 1,2-Dichlorobenzene-d4	12.192	152	91345	52.059	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.120%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071122\
Data File : VU049675.D
Acq On : 11 Jul 2022 14:58
Operator : SY/MD
Sample : N3639-01 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
COL62

Quant Time: Jul 12 02:34:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jul 12 02:23:25 2022
Response via : Initial Calibration

