

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071122\
 Data File : VU049681.D
 Acq On : 11 Jul 2022 17:19
 Operator : SY/MD
 Sample : N3625-06 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0ZC2

Quant Time: Jul 12 02:35:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jul 12 02:23:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	237042	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	220475	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	95113	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	95955	47.977	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.960%
7) Chloroethane-d5	1.906	69	70542	48.794	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.580%
11) 1,1-Dichloroethene-d2	2.565	63	123147	34.858	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.720%
21) 2-Butanone-d5	4.620	46	176629	103.451	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.450%
24) Chloroform-d	5.063	84	166480	46.997	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.000%
26) 1,2-Dichloroethane-d4	5.700	65	116493	48.590	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.180%
32) Benzene-d6	5.726	84	329928	49.258	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.520%
36) 1,2-Dichloropropane-d6	6.690	67	114203	50.691	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.380%
41) Toluene-d8	7.896	98	281469	48.255	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.520%
43) trans-1,3-Dichloroprop...	8.179	79	44120	46.481	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.960%
47) 2-Hexanone-d5	8.632	63	104026	102.649	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.650%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	155126	49.210	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.420%
66) 1,2-Dichlorobenzene-d4	12.195	152	90734	50.596	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071122\
Data File : VU049681.D
Acq On : 11 Jul 2022 17:19
Operator : SY/MD
Sample : N3625-06 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0ZC2

Quant Time: Jul 12 02:35:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jul 12 02:23:25 2022
Response via : Initial Calibration

