

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071122\
 Data File : VU049682.D
 Acq On : 11 Jul 2022 17:43
 Operator : SY/MD
 Sample : N3625-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Jul 12 02:35:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jul 12 02:23:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	221554	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	203251	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	88680	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	81974	43.852	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.700%
7) Chloroethane-d5	1.899	69	63217	46.785	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.560%
11) 1,1-Dichloroethene-d2	2.562	63	105352	31.906	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.820%
21) 2-Butanone-d5	4.620	46	174344	109.252	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.250%
24) Chloroform-d	5.063	84	157512	47.573	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.140%
26) 1,2-Dichloroethane-d4	5.700	65	110557	49.338	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.680%
32) Benzene-d6	5.726	84	302469	48.985	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.980%
36) 1,2-Dichloropropane-d6	6.690	67	105974	51.025	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.040%
41) Toluene-d8	7.896	98	250738	46.630	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.260%
43) trans-1,3-Dichloroprop...	8.179	79	42343	48.389	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.780%
47) 2-Hexanone-d5	8.632	63	98535	105.470	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.470%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	147108	50.621	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.240%
66) 1,2-Dichlorobenzene-d4	12.192	152	85730	51.274	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.540%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071122\
Data File : VU049682.D
Acq On : 11 Jul 2022 17:43
Operator : SY/MD
Sample : N3625-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Jul 12 02:35:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM071122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jul 12 02:23:25 2022
Response via : Initial Calibration

