

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071124\
 Data File : VU059896.D
 Acq On : 11 Jul 2024 09:21
 Operator : MD/SY
 Sample : VU0711MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK091

Quant Time: Jul 12 02:03:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 12 02:03:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	230271	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	236090	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	95309	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	77904	52.691	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 105.380%			
7) Chloroethane-d5	1.907	69	67110	53.964	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 107.920%			
11) 1,1-Dichloroethene-d2	2.560	63	96426	36.711	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 73.420%			
21) 2-Butanone-d5	4.615	46	101898	99.171	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 99.170%			
24) Chloroform-d	5.052	84	147607	51.554	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 103.100%			
26) 1,2-Dichloroethane-d4	5.692	65	90693	53.274	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 106.540%			
32) Benzene-d6	5.717	84	298591	49.390	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 98.780%			
36) 1,2-Dichloropropane-d6	6.682	67	96983	51.832	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 103.660%			
41) Toluene-d8	7.891	98	258267	44.922	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 89.840%			
43) trans-1,3-Dichloroprop...	8.174	79	40081	45.778	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 91.560%			
47) 2-Hexanone-d5	8.627	63	66674	84.552	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 84.550%			
56) 1,1,2,2-Tetrachloroeth...	10.749	84	147950	51.517	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 103.040%			
66) 1,2-Dichlorobenzene-d4	12.187	152	92754	52.960	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 105.920%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU071124\
Data File : VU059896.D
Acq On : 11 Jul 2024 09:21
Operator : MD/SY
Sample : VU0711MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK091

Quant Time: Jul 12 02:03:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 12 02:03:36 2024
Response via : Initial Calibration

